

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050523\
 Data File : VY013630.D
 Acq On : 05 May 2023 17:05
 Operator : KP/MD
 Sample : 02588-02
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ZONE-B-3-6-05012023

Quant Time: May 08 00:25:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	221693	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	375443	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	368036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	169819	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	163111	67.117	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.240%
35) Dibromofluoromethane	7.716	113	130848	55.412	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.820%
50) Toluene-d8	10.179	98	455231	54.192	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.380%
62) 4-Bromofluorobenzene	12.477	95	154134	54.427	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	108.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050523\
Data File : VY013630.D
Acq On : 05 May 2023 17:05
Operator : KP/MD
Sample : 02588-02
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ZONE-B-3-6-05012023

Quant Time: May 08 00:25:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

