

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050523\
 Data File : VY013631.D
 Acq On : 05 May 2023 17:28
 Operator : KP/MD
 Sample : 02588-03
 Misc : 6.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ZONE-B-6-9-05012023

Quant Time: May 08 00:25:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	214983	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	362185	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	359985	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	169433	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	160841	68.249	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.500%
35) Dibromofluoromethane	7.722	113	126279	55.435	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.860%
50) Toluene-d8	10.179	98	442469	54.601	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.200%
62) 4-Bromofluorobenzene	12.477	95	151382	55.412	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	110.820%
Target Compounds						
						Qvalue
16) Acetone	3.954	43	39827	42.913	ug/l	97
25) 2-Butanone	6.990	43	9286	7.889	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050523\
Data File : VY013631.D
Acq On : 05 May 2023 17:28
Operator : KP/MD
Sample : 02588-03
Misc : 6.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ZONE-B-6-9-05012023

Quant Time: May 08 00:25:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

