

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050523\
 Data File : VY013633.D
 Acq On : 05 May 2023 18:15
 Operator : KP/MD
 Sample : 02588-07
 Misc : 4.17g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ZONE-A-0-3-05022023

Quant Time: May 08 00:25:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	209282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	356593	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	343767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	157163	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	150809	65.735	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.480%
35) Dibromofluoromethane	7.716	113	120769	53.847	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.700%
50) Toluene-d8	10.179	98	431724	54.110	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.220%
62) 4-Bromofluorobenzene	12.477	95	141528	52.617	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	105.240%
Target Compounds						
16) Acetone	3.954	43	25868	28.632	ug/l	95
25) 2-Butanone	6.990	43	9068	7.914	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050523\
Data File : VY013633.D
Acq On : 05 May 2023 18:15
Operator : KP/MD
Sample : 02588-07
Misc : 4.17g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ZONE-A-0-3-05022023

Quant Time: May 08 00:25:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

