

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050523\
 Data File : VY013635.D
 Acq On : 05 May 2023 19:02
 Operator : KP/MD
 Sample : 02588-12
 Misc : 4.21g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ZONE-D-6-9-05022023

Quant Time: May 08 00:26:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	211167	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	355971	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	348207	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	164809	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	152920	66.060	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 132.120%	
35) Dibromofluoromethane	7.716	113	119940	53.571	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.140%	
50) Toluene-d8	10.179	98	432865	54.348	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 108.700%	
62) 4-Bromofluorobenzene	12.477	95	148896	55.453	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 110.900%	
Target Compounds						
16) Acetone	3.966	43	24188	26.533	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050523\
Data File : VY013635.D
Acq On : 05 May 2023 19:02
Operator : KP/MD
Sample : 02588-12
Misc : 4.21g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ZONE-D-6-9-05022023

Quant Time: May 08 00:26:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

