

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050624\
 Data File : VY018156.D
 Acq On : 06 May 2024 12:12
 Operator : SY/MD
 Sample : P2375-03
 Misc : 4.89g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-A-GRAB

Quant Time: May 06 23:45:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 16:21:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	242014	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	429295	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	358669	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	116777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	97553	51.138	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.280%
35) Dibromofluoromethane	7.716	113	110696	44.870	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.740%
50) Toluene-d8	10.179	98	470688	48.872	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.740%
62) 4-Bromofluorobenzene	12.483	95	129612	39.105	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.200%
Target Compounds						
					Qvalue	
16) Acetone	3.936	43	8889	24.460	ug/l	98
17) Carbon Disulfide	4.186	76	8832	1.278	ug/l	97
20) Methylene Chloride	4.710	84	30335	11.912	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050624\
Data File : VY018156.D
Acq On : 06 May 2024 12:12
Operator : SY/MD
Sample : P2375-03
Misc : 4.89g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-A-GRAB

Quant Time: May 06 23:45:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
Quant Title : SW846 8260
QLast Update : Wed May 01 16:21:20 2024
Response via : Initial Calibration

