

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050624\
 Data File : VY018160.D
 Acq On : 06 May 2024 13:47
 Operator : SY/MD
 Sample : P2386-03
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-050324-B

Quant Time: May 06 23:47:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 16:21:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	234361	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	429956	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	381401	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	147044	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	101973	55.201	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.400%
35) Dibromofluoromethane	7.716	113	123252	49.883	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.760%
50) Toluene-d8	10.179	98	475741	49.320	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.640%
62) 4-Bromofluorobenzene	12.483	95	147414	44.408	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.820%
Target Compounds						
16) Acetone	3.936	43	4994	14.191	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050624\
Data File : VY018160.D
Acq On : 06 May 2024 13:47
Operator : SY/MD
Sample : P2386-03
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-050324-B

Quant Time: May 06 23:47:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
Quant Title : SW846 8260
QLast Update : Wed May 01 16:21:20 2024
Response via : Initial Calibration

