

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050624\
 Data File : VY018154.D
 Acq On : 06 May 2024 11:06
 Operator : SY/MD
 Sample : VY0506SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0506SBS01

Quant Time: May 06 23:43:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 16:21:20 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Romaben Patel 05/07/2024
 Supervised By :Semsettin Yesilyurt 05/07/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	255956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	428377	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	371338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	172513	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	99247	49.192	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	98.380%		
35) Dibromofluoromethane	7.716	113	115186	46.790	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	93.580%		
50) Toluene-d8	10.179	98	451843	47.015	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	94.040%		
62) 4-Bromofluorobenzene	12.483	95	146748	44.370	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	88.740%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.900	85	43401	21.242	ug/l	98
3) Chloromethane	2.107	50	56965	21.339	ug/l	99
4) Vinyl Chloride	2.247	62	67158	23.251	ug/l	95
5) Bromomethane	2.637	94	47423	26.600	ug/l	97
6) Chloroethane	2.784	64	43694	24.291	ug/l	100
7) Trichlorofluoromethane	3.113	101	82259	21.392	ug/l	95
8) Diethyl Ether	3.521	74	27510	22.281	ug/l	98
9) 1,1,2-Trichlorotrifluo...	3.887	101	51752	21.566	ug/l	98
10) Methyl Iodide	4.088	142	63073	19.831	ug/l	97
11) Tert butyl alcohol	4.954	59	20713	125.966	ug/l #	78
12) 1,1-Dichloroethene	3.863	96	51173	20.912	ug/l	95
13) Acrolein	3.723	56	25500	104.192	ug/l	98
14) Allyl chloride	4.466	41	67003	20.710	ug/l	98
15) Acrylonitrile	5.143	53	56985	122.875	ug/l	98
16) Acetone	3.936	43	41820	108.807	ug/l	96
17) Carbon Disulfide	4.186	76	136294	18.653	ug/l	97
18) Methyl Acetate	4.472	43	22569	22.928	ug/l	99
19) Methyl tert-butyl Ether	5.210	73	126058	22.449	ug/l	98
20) Methylene Chloride	4.704	84	56215	20.872	ug/l	99
21) trans-1,2-Dichloroethene	5.210	96	55501	21.336	ug/l	97
22) Diisopropyl ether	6.112	45	147678	22.163	ug/l	97
23) Vinyl Acetate	6.051	43	463203	109.602	ug/l	98
24) 1,1-Dichloroethane	6.009	63	89089	22.142	ug/l	98
25) 2-Butanone	6.978	43	68494	114.549	ug/l	97
26) 2,2-Dichloropropane	6.966	77	80411	21.369	ug/l	98
27) cis-1,2-Dichloroethene	6.978	96	68362	22.036	ug/l	95
28) Bromochloromethane	7.332	49	34159	21.008	ug/l	89
29) Tetrahydrofuran	7.344	42	45683	120.178	ug/l	96
30) Chloroform	7.502	83	94285	22.144	ug/l	96
31) Cyclohexane	7.777	56	78113	20.206	ug/l	97
32) 1,1,1-Trichloroethane	7.704	97	82023	21.106	ug/l	98
36) 1,1-Dichloropropene	7.911	75	66544	20.250	ug/l	96
37) Ethyl Acetate	7.063	43	32410	22.664	ug/l #	95
38) Carbon Tetrachloride	7.893	117	72606	19.467	ug/l	95
39) Methylcyclohexane	9.179	83	93474	19.367	ug/l	98
40) Benzene	8.161	78	223265	21.144	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050624\
 Data File : VY018154.D
 Acq On : 06 May 2024 11:06
 Operator : SY/MD
 Sample : VY0506SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0506SBS01

Quant Time: May 06 23:43:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 16:21:20 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Romaben Patel 05/07/2024
 Supervised By :Semsettin Yesilyurt 05/07/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.313	41	18605m	22.062	ug/l	
42) 1,2-Dichloroethane	8.234	62	50173	20.510	ug/l	95
43) Isopropyl Acetate	8.271	43	63735	21.214	ug/l	95
44) Trichloroethene	8.935	130	61794	20.636	ug/l	94
45) 1,2-Dichloropropane	9.209	63	49363	21.426	ug/l	97
46) Dibromomethane	9.307	93	28165	20.108	ug/l	95
47) Bromodichloromethane	9.496	83	71179	20.699	ug/l	97
48) Methyl methacrylate	9.289	41	26968	19.609	ug/l	95
49) 1,4-Dioxane	9.295	88	7592	488.537	ug/l #	93
51) 4-Methyl-2-Pentanone	10.069	43	158909	108.298	ug/l	97
52) Toluene	10.240	92	143375	20.516	ug/l	98
53) t-1,3-Dichloropropene	10.465	75	67361	20.274	ug/l	99
54) cis-1,3-Dichloropropene	9.929	75	81973	21.001	ug/l	96
55) 1,1,2-Trichloroethane	10.642	97	40904	21.712	ug/l	95
56) Ethyl methacrylate	10.508	69	53728	20.171	ug/l #	90
57) 1,3-Dichloropropane	10.788	76	65416	21.735	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.782	63	119332	95.972	ug/l	93
59) 2-Hexanone	10.831	43	111307	106.283	ug/l	97
60) Dibromochloromethane	10.983	129	52536	20.531	ug/l	99
61) 1,2-Dibromoethane	11.087	107	37957	21.025	ug/l	100
64) Tetrachloroethene	10.721	164	52982	19.711	ug/l	99
65) Chlorobenzene	11.514	112	163603	20.874	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.593	131	53554	20.312	ug/l	98
67) Ethyl Benzene	11.593	91	269237	20.484	ug/l	96
68) m/p-Xylenes	11.703	106	215687	40.954	ug/l	97
69) o-Xylene	12.032	106	105873	20.926	ug/l	93
70) Styrene	12.044	104	175036	20.555	ug/l	98
71) Bromoform	12.209	173	30750	19.902	ug/l #	96
73) Isopropylbenzene	12.331	105	265506	20.449	ug/l	98
74) N-amyl acetate	12.142	43	57087	20.033	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.581	83	44836	21.702	ug/l	98
76) 1,2,3-Trichloropropane	12.629	75	30986m	21.402	ug/l	
77) Bromobenzene	12.611	156	63851	20.696	ug/l	93
78) n-propylbenzene	12.672	91	313137	20.944	ug/l	96
79) 2-Chlorotoluene	12.758	91	169127	19.917	ug/l	97
80) 1,3,5-Trimethylbenzene	12.812	105	213111	20.365	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.380	75	16122	20.286	ug/l	96
82) 4-Chlorotoluene	12.855	91	175591	20.180	ug/l	96
83) tert-Butylbenzene	13.075	119	194268	20.278	ug/l	100
84) 1,2,4-Trimethylbenzene	13.123	105	210176	20.503	ug/l	99
85) sec-Butylbenzene	13.251	105	288178	20.961	ug/l	97
86) p-Isopropyltoluene	13.367	119	236996	20.685	ug/l	98
87) 1,3-Dichlorobenzene	13.367	146	122064	20.686	ug/l	99
88) 1,4-Dichlorobenzene	13.446	146	120412	20.704	ug/l	99
89) n-Butylbenzene	13.696	91	217141	21.416	ug/l	98
90) Hexachloroethane	13.965	117	47448	20.384	ug/l	100
91) 1,2-Dichlorobenzene	13.739	146	108022	21.009	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.355	75	6234	20.003	ug/l	98
93) 1,2,4-Trichlorobenzene	15.007	180	61871	21.530	ug/l	99
94) Hexachlorobutadiene	15.111	225	34065	21.415	ug/l	99
95) Naphthalene	15.239	128	110470	20.878	ug/l	99
96) 1,2,3-Trichlorobenzene	15.422	180	50972	21.257	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050624\
Data File : VY018154.D
Acq On : 06 May 2024 11:06
Operator : SY/MD
Sample : VY0506SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024

Quant Time: May 06 23:43:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
Quant Title : SW846 8260
QLast Update : Wed May 01 16:21:20 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050624\
Data File : VY018154.D
Acq On : 06 May 2024 11:06
Operator : SY/MD
Sample : VY0506SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1


Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024


Quant Time: May 06 23:43:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
Quant Title : SW846 8260
QLast Update : Wed May 01 16:21:20 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 97
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion:168 Resp: 255956
Ion Ratio Lower Upper
168 100
99 53.2 39.1 58.7




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#2
Dichlorodifluoromethane
Concen: 21.242 ug/l
RT: 1.900 min Scan# 13
Delta R.T. -0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion: 85 Resp: 43401
Ion Ratio Lower Upper
85 100
87 32.9 16.0 48.0









#3
Chloromethane
Concen: 21.339 ug/l
RT: 2.107 min Scan# 47
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion: 50 Resp: 56965
Ion Ratio Lower Upper
50 100
52 32.3 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#4
Vinyl Chloride
Concen: 23.251 ug/l
RT: 2.247 min Scan# 70
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion: 62 Resp: 67158
Ion Ratio Lower Upper
62 100
64 29.9 26.1 39.1





#5
Bromomethane
Concen: 26.600 ug/l
RT: 2.637 min Scan# 134
Delta R.T. -0.012 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument : MSVOA_Y
ClientSampleId : VY0506SBS01



Tgt Ion: 94 Resp: 47423
Ion Ratio Lower Upper
94 100
96 95.5 74.1 111.1




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#6
Chloroethane
Concen: 24.291 ug/l
RT: 2.784 min Scan# 158
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion: 64 Resp: 43694
Ion Ratio Lower Upper
64 100
66 32.4 26.1 39.1





#7
 Trichlorofluoromethane
 Concen: 21.392 ug/l
 RT: 3.113 min Scan# 212
 Delta R.T. -0.012 min
 Lab File: VY018154.D
 Acq: 06 May 2024 11:06




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0506SBS01



Tgt Ion:101 Resp: 82259
 Ion Ratio Lower Upper
 101 100
 103 57.8 49.3 73.9




Manual Integrations
 APPROVED


Reviewed By :Romaben Patel 05/07/2024
 Supervised By :Semsettin Yesilyurt 05/07/2024



#8
 Diethyl Ether
 Concen: 22.281 ug/l
 RT: 3.521 min Scan# 279
 Delta R.T. -0.006 min
 Lab File: VY018154.D
 Acq: 06 May 2024 11:06

Tgt Ion: 74 Resp: 27510
 Ion Ratio Lower Upper
 74 100
 45 81.0 41.3 124.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.566 ug/l
RT: 3.887 min Scan# 339
Delta R.T. -0.012 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion:	101	Resp:	51752
Ion Ratio	Lower	Upper	
101	100		
85	43.4	34.5	51.7
151	79.9	61.7	92.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#10
Methyl Iodide
Concen: 19.831 ug/l
RT: 4.088 min Scan# 372
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion:	142	Resp:	63073
Ion Ratio	Lower	Upper	
142	100		
127	39.0	32.8	49.2
141	14.0	11.4	17.2





#11

Tert butyl alcohol

Concen: 125.966 ug/l

RT: 4.954 min Scan# 51

Delta R.T. 0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :

VY0506SBS01

Tgt Ion: 59 Resp: 20713

Ion Ratio Lower Upper

59 100





57 10.9 3.0 4.6#

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024

Supervised By :Semsettin Yesilyurt 05/07/2024



#12

1,1-Dichloroethene

Concen: 20.912 ug/l

RT: 3.863 min Scan# 335

Delta R.T. -0.012 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06




Tgt Ion: 96 Resp: 51173


Ion Ratio Lower Upper

96 100

61 135.7 114.6 171.8

98 62.8 50.8 76.2





#13

Acrolein

Concen: 104.192 ug/l

RT: 3.723 min Scan# 311

Delta R.T. -0.006 min

Lab File: VY018154.D


Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :

VY0506SBS01



Tgt Ion: 56 Resp: 25500

Ion Ratio Lower Upper

56 100




55 67.6 55.7 83.5

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024

Supervised By :Semsettin Yesilyurt 05/07/2024



#14

Allyl chloride

Concen: 20.710 ug/l

RT: 4.466 min Scan# 434

Delta R.T. -0.012 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06




Tgt Ion: 41 Resp: 67003


Ion Ratio Lower Upper

41 100

39 74.8 58.3 87.5


76 48.4 39.0 58.6





#15
Acrylonitrile
Concen: 122.875 ug/l
RT: 5.143 min Scan# 546
Delta R.T. -0.018 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01




Tgt Ion: 53 Resp: 56985
Ion Ratio Lower Upper
53 100
52 79.0 64.7 97.1
51 36.1 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024




#16
Acetone
Concen: 108.807 ug/l
RT: 3.936 min Scan# 347
Delta R.T. -0.012 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Tgt Ion: 43 Resp: 41820
Ion Ratio Lower Upper
43 100
58 36.8 31.7 47.5





#17
Carbon Disulfide
Concen: 18.653 ug/l
RT: 4.186 min Scan# 387
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion: 76 Resp: 136294
Ion Ratio Lower Upper
76 100
78 9.5 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#18
Methyl Acetate
Concen: 22.928 ug/l
RT: 4.472 min Scan# 435
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion: 43 Resp: 22569
Ion Ratio Lower Upper
43 100
74 29.6 23.3 34.9





#19
Methyl tert-butyl Ether
Concen: 22.449 ug/l
RT: 5.210 min Scan# 556
Delta R.T. -0.012 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion: 73 Resp: 126058
Ion Ratio Lower Upper
73 100
57 20.6 17.3 25.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#20
Methylene Chloride
Concen: 20.872 ug/l
RT: 4.704 min Scan# 473
Delta R.T. -0.012 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion: 84 Resp: 56215
Ion Ratio Lower Upper
84 100
49 104.9 84.4 126.6
51 33.5 26.0 39.0
86 62.1 49.5 74.3





#21
trans-1,2-Dichloroethene
Concen: 21.336 ug/l
RT: 5.210 min Scan# 556
Delta R.T. -0.012 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion	Ratio	Lower	Upper
96	100		
61	119.4	97.5	146.3
98	66.2	50.5	75.7




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#22
Diisopropyl ether
Concen: 22.163 ug/l
RT: 6.112 min Scan# 704
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion	Ratio	Lower	Upper
45	100		
43	54.3	42.2	63.2
87	32.2	25.0	37.6
59	10.9	10.5	15.7





#23

Vinyl Acetate

Concen: 109.602 ug/l

RT: 6.051 min Scan# 69

Delta R.T. -0.013 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :

VY0506SBS01

Tgt Ion: 43 Resp: 463203

Ion Ratio Lower Upper

43 100





86 14.1 10.6 16.0

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024

Supervised By :Semsettin Yesilyurt 05/07/2024



#24

1,1-Dichloroethane

Concen: 22.142 ug/l

RT: 6.009 min Scan# 687

Delta R.T. -0.012 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06




Tgt Ion: 63 Resp: 89089


Ion Ratio Lower Upper

63 100

98 9.0 4.0 12.2


100 5.3 2.6 8.0





#25
2-Butanone
Concen: 114.549 ug/l
RT: 6.978 min Scan# 846
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion: 43 Resp: 68494
Ion Ratio Lower Upper
43 100
72 32.2 27.3 40.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#26
2,2-Dichloropropane
Concen: 21.369 ug/l
RT: 6.966 min Scan# 844
Delta R.T. -0.012 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion: 77 Resp: 80411
Ion Ratio Lower Upper
77 100
97 24.9 12.0 36.1





#27
cis-1,2-Dichloroethene
Concen: 22.036 ug/l
RT: 6.978 min Scan# 846
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01




Tgt Ion: 96 Resp: 68362
Ion Ratio Lower Upper
96 100
61 120.5 0.0 256.2
98 64.4 0.0 130.0

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024




#28
Bromochloromethane
Concen: 21.008 ug/l
RT: 7.332 min Scan# 904
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06



Tgt Ion: 49 Resp: 34159
Ion Ratio Lower Upper
49 100
129 3.1 0.0 4.6
130 101.7 72.7 109.1









#29
Tetrahydrofuran
Concen: 120.178 ug/l
RT: 7.344 min Scan# 905
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion:	42	Resp:	45683
Ion	Ratio	Lower	Upper
42	100		
72	50.6	43.0	64.4
71	48.0	40.8	61.2




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#30
Chloroform
Concen: 22.144 ug/l
RT: 7.502 min Scan# 932
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion:	83	Resp:	94285
Ion	Ratio	Lower	Upper
83	100		
85	68.4	52.2	78.4









#31
Cyclohexane
Concen: 20.206 ug/l
RT: 7.777 min Scan# 977
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion: 56 Resp: 78113
Ion Ratio Lower Upper
56 100
69 38.9 28.7 43.1
84 95.0 77.7 116.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#32
1,1,1-Trichloroethane
Concen: 21.106 ug/l
RT: 7.704 min Scan# 965
Delta R.T. 0.001 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion: 97 Resp: 82023
Ion Ratio Lower Upper
97 100
99 65.2 50.3 75.5
61 40.2 32.6 48.8






#33
 1,2-Dichloroethane-d4
 Concen: 49.192 ug/l
 RT: 8.136 min Scan# 1036
 Delta R.T. -0.006 min
 Lab File: VY018154.D
 Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :




VY0506SBS01



Tgt Ion:	65	Resp:	99247
Ion	Ratio	Lower	Upper
65	100		
67	54.5	0.0	109.6




Manual Integrations
 APPROVED


Reviewed By :Romaben Patel 05/07/2024
 Supervised By :Semsettin Yesilyurt 05/07/2024



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.685 min Scan# 1126
 Delta R.T. -0.006 min
 Lab File: VY018154.D
 Acq: 06 May 2024 11:06


Tgt Ion:	114	Resp:	428377
Ion	Ratio	Lower	Upper
114	100		
63	16.6	0.0	35.0
88	14.3	0.0	27.2





#35
Dibromofluoromethane
Concen: 46.790 ug/l
RT: 7.716 min Scan# 96
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion:113 Resp: 115186
Ion Ratio Lower Upper
113 100
111 103.5 82.2 123.4
192 21.3 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#36
1,1-Dichloropropene
Concen: 20.250 ug/l
RT: 7.911 min Scan# 999
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion: 75 Resp: 66544
Ion Ratio Lower Upper
75 100
110 41.9 19.0 57.0
77 32.3 25.3 37.9









#37
Ethyl Acetate
Concen: 22.664 ug/l
RT: 7.063 min Scan# 860
Delta R.T. -0.013 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion: 43 Resp: 32410
Ion Ratio Lower Upper
43 100
61 14.8 12.6 19.0
70 13.4 7.8 11.8#




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#38
Carbon Tetrachloride
Concen: 19.467 ug/l
RT: 7.893 min Scan# 996
Delta R.T. -0.012 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion:117 Resp: 72606
Ion Ratio Lower Upper
117 100
119 90.1 75.8 113.6
121 27.6 24.0 36.0









#39
Methylcyclohexane
Concen: 19.367 ug/l
RT: 9.179 min Scan# 1207
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion:	83	Resp:	93474
Ion	Ratio	Lower	Upper
83	100		
55	64.9	51.5	77.3
98	53.0	40.5	60.7




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#40
Benzene
Concen: 21.144 ug/l
RT: 8.161 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion:	78	Resp:	223265
Ion	Ratio	Lower	Upper
78	100		
77	21.3	19.5	29.3





#41
Methacrylonitrile
Concen: 22.062 ug/l m
RT: 7.313 min Scan# 90
Delta R.T. -0.019 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion: 41 Resp: 18605
Ion Ratio Lower Upper
41 100
39 57.9 48.0 72.0
67 0.0 75.1 112.7#
52 0.0 29.8 44.6#




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#42
1,2-Dichloroethane
Concen: 20.510 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion: 62 Resp: 50173
Ion Ratio Lower Upper
62 100
98 12.8 0.0 22.0









#43
Isopropyl Acetate
Concen: 21.214 ug/l
RT: 8.271 min Scan# 1057
Delta R.T. 0.001 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion: 43 Resp: 63735
Ion Ratio Lower Upper
43 100
61 31.1 27.9 41.9
87 16.5 13.1 19.7




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#44
Trichloroethene
Concen: 20.636 ug/l
RT: 8.935 min Scan# 1167
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion:130 Resp: 61794
Ion Ratio Lower Upper
130 100
95 88.1 0.0 188.0





#45
1,2-Dichloropropane
Concen: 21.426 ug/l
RT: 9.209 min Scan# 1212
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion: 63 Resp: 49363
Ion Ratio Lower Upper
63 100
65 29.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#46
Dibromomethane
Concen: 20.108 ug/l
RT: 9.307 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion: 93 Resp: 28165
Ion Ratio Lower Upper
93 100
95 82.6 66.2 99.4
174 109.5 80.9 121.3









#47
Bromodichloromethane
Concen: 20.699 ug/l
RT: 9.496 min Scan# 1258
Delta R.T. -0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion:	83	Resp:	71179
Ion	Ratio	Lower	Upper
83	100		
85	64.7	50.1	75.1
127	8.5	7.4	11.0




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#48
Methyl methacrylate
Concen: 19.609 ug/l
RT: 9.289 min Scan# 1225
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion:	41	Resp:	26968
Ion	Ratio	Lower	Upper
41	100		
69	102.6	84.1	126.1
39	60.7	43.6	65.4





#49
1,4-Dioxane
Concen: 488.537 ug/l
RT: 9.295 min Scan# 1225
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion: 88 Resp: 7592
Ion Ratio Lower Upper
88 100
43 29.1 0.0 0.0#
58 55.9 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#50
Toluene-d8
Concen: 47.015 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion: 98 Resp: 451843
Ion Ratio Lower Upper
98 100
100 66.6 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 108.298 ug/l
RT: 10.069 min Scan# 1352
Delta R.T. -0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion: 43 Resp: 158909
Ion Ratio Lower Upper
43 100
58 42.7 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#52
Toluene
Concen: 20.516 ug/l
RT: 10.240 min Scan# 1381
Delta R.T. -0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion: 92 Resp: 143375
Ion Ratio Lower Upper
92 100
91 170.8 139.0 208.6





#53

t-1,3-Dichloropropene

Concen: 20.274 ug/l

RT: 10.465 min Scan# 1417

Delta R.T. 0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :

VY0506SBS01

Tgt Ion: 75 Resp: 67361

Ion Ratio Lower Upper





75 100

77 30.8 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#54

cis-1,3-Dichloropropene

Concen: 21.001 ug/l

RT: 9.929 min Scan# 1330

Delta R.T. -0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06




Tgt Ion: 75 Resp: 81973


Ion Ratio Lower Upper

75 100

77 30.5 25.4 38.0

39 40.3 29.8 44.6





#55

1,1,2-Trichloroethane

Concen: 21.712 ug/l

RT: 10.642 min Scan# 1447

Delta R.T. 0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :





VY0506SBS01

Tgt Ion:	97	Resp:	40904
Ion	Ratio	Lower	Upper
97	100		
83	80.1	67.5	101.3
85	49.0	42.9	64.3
99	56.0	46.6	70.0

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#56

Ethyl methacrylate

Concen: 20.171 ug/l




RT: 10.508 min Scan# 1425


Delta R.T. -0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Tgt Ion:	69	Resp:	53728
Ion	Ratio	Lower	Upper
69	100		
41	59.6	42.3	63.5
39	34.6	22.9	34.3#





#57

1,3-Dichloropropane

Concen: 21.735 ug/l

RT: 10.788 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :

VY0506SBS01

Tgt Ion: 76 Resp: 65416

Ion Ratio Lower Upper

76 100





78 32.5 26.2 39.4

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024

Supervised By :Semsettin Yesilyurt 05/07/2024



#58

2-Chloroethyl Vinyl ether

Concen: 95.972 ug/l

RT: 9.782 min Scan# 1306

Delta R.T. 0.000 min

Lab File: VY018154.D




Acq: 06 May 2024 11:06


Tgt Ion: 63 Resp: 119332

Ion Ratio Lower Upper

63 100


106 33.3 23.7 35.5





#59
2-Hexanone
Concen: 106.283 ug/l
RT: 10.831 min Scan# 1477
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion: 43 Resp: 111307
Ion Ratio Lower Upper
43 100
58 60.4 31.4 94.0




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#60
Dibromochloromethane
Concen: 20.531 ug/l
RT: 10.983 min Scan# 1503
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion:129 Resp: 52536
Ion Ratio Lower Upper
129 100
127 78.6 39.0 117.0





#61
1,2-Dibromoethane
Concen: 21.025 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01




Tgt Ion:107 Resp: 37957
Ion Ratio Lower Upper
107 100
109 94.8 75.9 113.9

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024




#62
4-Bromofluorobenzene
Concen: 44.370 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Tgt Ion: 95 Resp: 146748
Ion Ratio Lower Upper
95 100
174 89.5 0.0 175.6
176 86.3 0.0 171.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 1586
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion:117 Resp: 371338
Ion Ratio Lower Upper
117 100
82 50.0 42.4 63.6
119 32.4 25.9 38.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#64
Tetrachloroethene
Concen: 19.711 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion:164 Resp: 52982
Ion Ratio Lower Upper
164 100
166 127.5 102.6 154.0
129 90.9 72.0 108.0
131 85.5 70.0 105.0





#65

Chlorobenzene

Concen: 20.874 ug/l

RT: 11.514 min Scan# 1590

Delta R.T. -0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :

VY0506SBS01

Tgt Ion:112 Resp: 163603

Ion Ratio Lower Upper

112 100





114 30.7 26.4 39.6

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024

Supervised By :Semsettin Yesilyurt 05/07/2024



#66

1,1,1,2-Tetrachloroethane

Concen: 20.312 ug/l

RT: 11.593 min Scan# 1603

Delta R.T. 0.006 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06




Tgt Ion:131 Resp: 53554


Ion Ratio Lower Upper

131 100

133 93.2 47.8 143.3

119 66.3 34.1 102.2





#67

Ethyl Benzene

Concen: 20.484 ug/l

RT: 11.593 min Scan# 1603

Delta R.T. 0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :

VY0506SBS01

Tgt Ion: 91 Resp: 269237

Ion Ratio Lower Upper

91 100





106 32.9 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024

Supervised By :Semsettin Yesilyurt 05/07/2024



#68

m/p-Xylenes

Concen: 40.954 ug/l

RT: 11.703 min Scan# 1621

Delta R.T. -0.000 min

Lab File: VY018154.D




Acq: 06 May 2024 11:06


Tgt Ion:106 Resp: 215687

Ion Ratio Lower Upper

106 100

91 189.8 156.0 234.0





#69

o-Xylene

Concen: 20.926 ug/l

RT: 12.032 min Scan# 1675

Delta R.T. 0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :

VY0506SBS01

Tgt Ion:106 Resp: 105873

Ion Ratio Lower Upper

106 100





91 196.2 103.6 310.8

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024

Supervised By :Semsettin Yesilyurt 05/07/2024



#70

Styrene

Concen: 20.555 ug/l

RT: 12.044 min Scan# 1677

Delta R.T. 0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06




Tgt Ion:104 Resp: 175036


Ion Ratio Lower Upper

104 100

78 43.9 37.0 55.4

103 54.3 44.2 66.2





#71

Bromoform

Concen: 19.902 ug/l

RT: 12.209 min Scan# 1703

Delta R.T. -0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :





VY0506SBS01

Tgt Ion:	173	Resp:	30750
Ion Ratio	Lower	Upper	
173	100		
175	47.4	25.1	75.3
254	0.0	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.428 min Scan# 1904


Delta R.T. 0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Tgt Ion:	152	Resp:	172513
Ion Ratio	Lower	Upper	
152	100		
115	55.5	28.2	84.7
150	164.2	0.0	345.6





#73

Isopropylbenzene

Concen: 20.449 ug/l

RT: 12.331 min Scan# 1723

Delta R.T. -0.000 min

Lab File: VY018154.D


Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :

VY0506SBS01



Tgt Ion:105 Resp: 265506

Ion Ratio Lower Upper

105 100




120 27.9 13.4 40.1

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024

Supervised By :Semsettin Yesilyurt 05/07/2024



#74

N-ethyl acetate

Concen: 20.033 ug/l

RT: 12.142 min Scan# 1693

Delta R.T. -0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Tgt Ion: 43 Resp: 57087




Ion Ratio Lower Upper


43 100

70 46.7 41.4 62.2

55 25.4 23.3 34.9

61 27.3 22.6 33.8









#75
1,1,2,2-Tetrachloroethane
Concen: 21.702 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. -0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion	83	Resp	44836
Ion	Ratio	Lower	Upper
83	100		
131	11.1	5.0	14.9
85	62.9	32.0	96.0




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#76
1,2,3-Trichloropropane
Concen: 21.402 ug/l m
RT: 12.629 min Scan# 1773
Delta R.T. -0.007 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion	75	Resp	30986
Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0









#77
Bromobenzene
Concen: 20.696 ug/l
RT: 12.611 min Scan# 1769
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion	Ratio	Lower	Upper
156	100		
77	160.7	87.1	261.3
158	95.2	48.6	145.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#78
n-propylbenzene
Concen: 20.944 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion	Ratio	Lower	Upper
91	100		
120	24.6	11.5	34.4





#79
2-Chlorotoluene
Concen: 19.917 ug/l
RT: 12.758 min Scan# 1794
Delta R.T. 0.001 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion: 91 Resp: 169127
Ion Ratio Lower Upper
91 100
126 37.7 17.8 53.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#80
1,3,5-Trimethylbenzene
Concen: 20.365 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion:105 Resp: 213111
Ion Ratio Lower Upper
105 100
120 51.5 25.1 75.2





#81
trans-1,4-Dichloro-2-butene
Concen: 20.286 ug/l
RT: 12.380 min Scan# 1731
Delta R.T. 0.001 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion: 75 Resp: 16122
Ion Ratio Lower Upper
75 100
53 85.0 66.8 100.2
89 49.8 35.8 53.6




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#82
4-Chlorotoluene
Concen: 20.180 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion: 91 Resp: 175591
Ion Ratio Lower Upper
91 100
126 37.8 17.6 52.9









#83
tert-Butylbenzene
Concen: 20.278 ug/l
RT: 13.075 min Scan# 1846
Delta R.T. 0.001 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion:119 Resp: 194268
Ion Ratio Lower Upper
119 100
91 62.4 31.3 93.8
134 27.0 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#84
1,2,4-Trimethylbenzene
Concen: 20.503 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.006 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion:105 Resp: 210176
Ion Ratio Lower Upper
105 100
120 47.1 23.2 69.6





#85
sec-Butylbenzene
Concen: 20.961 ug/l
RT: 13.251 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion:105 Resp: 288178
Ion Ratio Lower Upper
105 100
134 21.3 10.1 30.1




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#86
p-Isopropyltoluene
Concen: 20.685 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion:119 Resp: 236996
Ion Ratio Lower Upper
119 100
134 27.7 13.1 39.3
91 22.0 10.7 32.1





#87
1,3-Dichlorobenzene
Concen: 20.686 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion:146 Resp: 122064
Ion Ratio Lower Upper
146 100
111 36.9 19.0 57.0
148 64.2 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#88
1,4-Dichlorobenzene
Concen: 20.704 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion:146 Resp: 120412
Ion Ratio Lower Upper
146 100
111 37.9 18.4 55.4
148 64.9 32.4 97.2





#89

n-Butylbenzene

Concen: 21.416 ug/l

RT: 13.696 min Scan# 1948

Delta R.T. 0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :





VY0506SBS01

Tgt Ion:	91	Resp:	217141
Ion	Ratio	Lower	Upper
91	100		
92	53.9	26.6	79.7
134	28.4	12.9	38.7

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#90

Hexachloroethane

Concen: 20.384 ug/l




RT: 13.965 min Scan# 1992


Delta R.T. 0.006 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Tgt Ion:	117	Resp:	47448
Ion	Ratio	Lower	Upper
117	100		
201	74.9	37.4	112.2









#91
1,2-Dichlorobenzene
Concen: 21.009 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01

Tgt Ion	Ratio	Lower	Upper
146	100		
111	38.7	19.9	59.7
148	62.6	31.8	95.4




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.003 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06


Tgt Ion	Ratio	Lower	Upper
75	100		
155	96.9	48.6	145.8
157	130.2	62.7	188.2





#93
1,2,4-Trichlorobenzene
Concen: 21.530 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06




Instrument :
MSVOA_Y
ClientSampleId :
VY0506SBS01



Tgt Ion:180 Resp: 61871
Ion Ratio Lower Upper
180 100
182 96.2 47.9 143.8
145 29.5 14.2 42.6




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#94
Hexachlorobutadiene
Concen: 21.415 ug/l
RT: 15.111 min Scan# 2180
Delta R.T. 0.000 min
Lab File: VY018154.D
Acq: 06 May 2024 11:06

Tgt Ion:225 Resp: 34065
Ion Ratio Lower Upper
225 100
223 60.3 30.4 91.2
227 63.2 31.9 95.9





#95

Naphthalene

Concen: 20.878 ug/l

RT: 15.239 min Scan# 2200

Delta R.T. 0.006 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Instrument :

MSVOA_Y

ClientSampleId :





VY0506SBS01

Tgt Ion:	128	Resp:	110470
Ion	Ratio	Lower	Upper
128	100		
127	12.7	10.6	16.0
129	11.2	8.6	13.0

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 05/07/2024
Supervised By :Semsettin Yesilyurt 05/07/2024



#96

1,2,3-Trichlorobenzene

Concen: 21.257 ug/l

RT: 15.422 min Scan# 2231

Delta R.T. 0.000 min

Lab File: VY018154.D

Acq: 06 May 2024 11:06

Tgt Ion:	180	Resp:	50972
Ion	Ratio	Lower	Upper
180	100		
182	96.8	47.6	142.9
145	32.4	15.4	46.1

