

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050824\
 Data File : VY018193.D
 Acq On : 08 May 2024 17:11
 Operator : SY/MD
 Sample : P2414-01
 Misc : 4.74g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 1011UMR-SS004-0212-01

Quant Time: May 09 02:28:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	109491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	198845	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	172887	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	65460	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	51843	64.362	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.720%
35) Dibromofluoromethane	7.710	113	56658	54.217	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.440%
50) Toluene-d8	10.179	98	219115	53.208	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.420%
62) 4-Bromofluorobenzene	12.477	95	67845	51.016	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	102.040%
Target Compounds						
20) Methylene Chloride	4.704	84	7622	6.709	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050824\
Data File : VY018193.D
Acq On : 08 May 2024 17:11
Operator : SY/MD
Sample : P2414-01
Misc : 4.74g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
1011UMR-SS004-0212-01

Quant Time: May 09 02:28:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

