

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050824\
 Data File : VY018198.D
 Acq On : 08 May 2024 19:12
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: May 09 02:31:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	176983	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	331228	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	296336	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	116356	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	79695	61.209	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.420%
35) Dibromofluoromethane	7.710	113	95742	55.000	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.000%
50) Toluene-d8	10.173	98	373553	54.456	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.920%
62) 4-Bromofluorobenzene	12.477	95	115710	52.233	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	104.460%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050824\
Data File : VY018198.D
Acq On : 08 May 2024 19:12
Operator : SY/MD
Sample : VIBLK
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: May 09 02:31:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

