

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050924\
 Data File : VY018210.D
 Acq On : 09 May 2024 15:18
 Operator : SY/MD
 Sample : P2408-04
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ETGI-285

Quant Time: May 10 01:18:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	162543	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	300730	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	257086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	95049	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	71499	59.793	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.580%
35) Dibromofluoromethane	7.709	113	84777	53.640	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.280%
50) Toluene-d8	10.178	98	335379	53.850	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.700%
62) 4-Bromofluorobenzene	12.483	95	98141	48.795	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050924\
Data File : VY018210.D
Acq On : 09 May 2024 15:18
Operator : SY/MD
Sample : P2408-04
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ETGI-285

Quant Time: May 10 01:18:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

