

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050924\
 Data File : VY018211.D
 Acq On : 09 May 2024 15:42
 Operator : SY/MD
 Sample : P2408-02
 Misc : 4.50g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BIN-1

Quant Time: May 10 01:18:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	180316	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	339834	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	295722	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	113830	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	83121	62.661	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 125.320%	
35) Dibromofluoromethane	7.716	113	96365	53.956	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.920%	
50) Toluene-d8	10.179	98	384016	54.564	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 109.120%	
62) 4-Bromofluorobenzene	12.483	95	114659	50.448	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 100.900%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.704	84	13294	7.105	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050924\
Data File : VY018211.D
Acq On : 09 May 2024 15:42
Operator : SY/MD
Sample : P2408-02
Misc : 4.50g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BIN-1

Quant Time: May 10 01:18:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

