



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050924\
Data File : VY018200.D
Acq On : 09 May 2024 09:43
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Title : SW846 8260
Last Update : Wed May 08 01:40:14 2024



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1738

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.0	25213	PASS
75	95	30	60	48.9	72600	PASS
95	95	100	100	100.0	148475	PASS
96	95	5	9	6.7	10006	PASS
173	174	0.00	2	0.9	1105	PASS
174	95	50	100	86.1	127899	PASS
175	174	5	9	7.2	9254	PASS
176	174	95	101	95.6	122288	PASS
177	176	5	9	6.7	8242	PASS