

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051021\
 Data File : VY004728.D
 Acq On : 10 May 2021 14:32
 Operator : SY/MD
 Sample : VY0510SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0510SBL01

Quant Time: May 11 04:18:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051021S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 11 04:13:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	156948	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	283275	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	273765	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	125534	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	108874	58.28	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.56%
35) Dibromofluoromethane	7.722	113	89103	57.46	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.92%
50) Toluene-d8	10.185	98	348938	49.81	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.62%
62) 4-Bromofluorobenzene	12.483	95	121026	51.19	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.38%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051021\
Data File : VY004728.D
Acq On : 10 May 2021 14:32
Operator : SY/MD
Sample : VY0510SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0510SBL01

Quant Time: May 11 04:18:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051021S.M
Quant Title : SW846 8260
QLast Update : Tue May 11 04:13:44 2021
Response via : Initial Calibration

