

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051024\
 Data File : VY018229.D
 Acq On : 10 May 2024 13:20
 Operator : SY/MD
 Sample : P2444-03
 Misc : 6.09g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-DD-VOC

Quant Time: May 11 01:00:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	170184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	318278	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	291198	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	112413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	92458	73.849	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	147.700%
35) Dibromofluoromethane	7.716	113	95585	57.144	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.280%
50) Toluene-d8	10.179	98	361319	54.816	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.640%
62) 4-Bromofluorobenzene	12.483	95	114650	53.860	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	107.720%
Target Compounds						
16) Acetone	3.948	43	10083	40.357	ug/l	93
20) Methylene Chloride	4.704	84	19719	11.166	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051024\
Data File : VY018229.D
Acq On : 10 May 2024 13:20
Operator : SY/MD
Sample : P2444-03
Misc : 6.09g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-DD-VOC

Quant Time: May 11 01:00:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

