

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051024\
 Data File : VY018232.D
 Acq On : 10 May 2024 14:32
 Operator : SY/MD
 Sample : P2432-03
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SW-01-011

Quant Time: May 11 01:01:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	163317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	315220	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	281730	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	105394	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	82634	68.777	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 137.560%			
35) Dibromofluoromethane	7.710	113	93430	56.398	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 112.800%			
50) Toluene-d8	10.179	98	360904	55.284	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 110.560%			
62) 4-Bromofluorobenzene	12.483	95	109414	51.899	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 103.800%			
Target Compounds						
20) Methylene Chloride	4.704	84	26341	15.543	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051024\
Data File : VY018232.D
Acq On : 10 May 2024 14:32
Operator : SY/MD
Sample : P2432-03
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SW-01-011

Quant Time: May 11 01:01:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

