

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051024\
 Data File : VY018234.D
 Acq On : 10 May 2024 15:20
 Operator : SY/MD
 Sample : P2437-01
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 359

Quant Time: May 11 01:02:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	171776	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	321869	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	284335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	104575	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	77873	61.623	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.240%
35) Dibromofluoromethane	7.715	113	91836	54.290	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.580%
50) Toluene-d8	10.178	98	365539	54.838	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.680%
62) 4-Bromofluorobenzene	12.483	95	108224	50.274	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	100.540%
Target Compounds					Qvalue	
20) Methylene Chloride	4.710	84	20432	11.463	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051024\
Data File : VY018234.D
Acq On : 10 May 2024 15:20
Operator : SY/MD
Sample : P2437-01
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
359

Quant Time: May 11 01:02:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

