

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051024\
 Data File : VY018225.D
 Acq On : 10 May 2024 11:37
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 11 00:47:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.397	0.351	11.6	73	0.00
3 P	Chloromethane	0.531	0.531	0.0	80	0.00
4 C	Vinyl Chloride	0.643	0.649	-0.9#	80	0.00
5 T	Bromomethane	0.447	0.436	2.5	81	0.00
6 T	Chloroethane	0.408	0.414	-1.5	81	0.00
7 T	Trichlorofluoromethane	0.754	0.732	2.9	79	0.00
8 T	Diethyl Ether	0.248	0.226	8.9	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.480	0.454	5.4	77	0.00
10 T	Methyl Iodide	0.561	0.519	7.5	69	0.00
11 T	Tert butyl alcohol	0.037	0.033	10.8	88	0.00
12 CM	1,1-Dichloroethene	0.484	0.454	6.2#	75	0.00
13 T	Acrolein	0.040	0.026	35.0#	73	0.00
14 T	Allyl chloride	0.615	0.592	3.7	78	0.00
15 T	Acrylonitrile	0.095	0.092	3.2	83	0.00
16 T	Acetone	0.073	0.085	-16.4	98	0.00
17 T	Carbon Disulfide	1.316	1.260	4.3	71	0.00
18 T	Methyl Acetate	0.209	0.185	11.5	81	0.00
19 T	Methyl tert-butyl Ether	1.077	1.001	7.1	79	0.00
20 T	Methylene Chloride	0.519	0.507	2.3	86	0.00
21 T	trans-1,2-Dichloroethene	0.513	0.483	5.8	75	0.00
22 T	Diisopropyl ether	1.305	1.225	6.1	78	0.00
23 T	Vinyl Acetate	0.793	0.771	2.8	82	0.00
24 P	1,1-Dichloroethane	0.804	0.766	4.7	77	0.00
25 T	2-Butanone	0.117	0.119	-1.7	88	0.00
26 T	2,2-Dichloropropane	0.727	0.696	4.3	78	0.00
27 T	cis-1,2-Dichloroethene	0.606	0.565	6.8	75	0.00
28 T	Bromochloromethane	0.321	0.332	-3.4	91	0.00
29 T	Tetrahydrofuran	0.077	0.075	2.6	87	0.00
30 C	Chloroform	0.839	0.791	5.7#	77	0.00
31 T	Cyclohexane	0.763	0.700	8.3	76	0.00
32 T	1,1,1-Trichloroethane	0.752	0.721	4.1	78	0.00
33 S	1,2-Dichloroethane-d4	0.368	0.362	1.6	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.263	0.268	-1.9	100	0.00
36 T	1,1-Dichloropropene	0.377	0.369	2.1	77	0.00
37 T	Ethyl Acetate	0.159	0.162	-1.9	85	0.00
38 T	Carbon Tetrachloride	0.412	0.410	0.5	78	0.00
39 T	Methylcyclohexane	0.543	0.538	0.9	76	0.00
40 TM	Benzene	1.212	1.175	3.1	76	0.00
41 T	Methacrylonitrile	0.092	0.084	8.7	72	-0.01
42 TM	1,2-Dichloroethane	0.264	0.266	-0.8	83	0.00
43 T	Isopropyl Acetate	0.325	0.324	0.3	86	0.00
44 TM	Trichloroethene	0.330	0.319	3.3	76	0.00
45 C	1,2-Dichloropropane	0.265	0.253	4.5#	76	0.00
46 T	Dibromomethane	0.152	0.146	3.9	80	0.00
47 T	Bromodichloromethane	0.376	0.366	2.7	78	0.00
48 T	Methyl methacrylate	0.149	0.145	2.7	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051024\
 Data File : VY018225.D
 Acq On : 10 May 2024 11:37
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 11 00:47:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	81	0.00
50 S	Toluene-d8	1.035	1.035	0.0	97	0.00
51 T	4-Methyl-2-Pentanone	0.163	0.167	-2.5	87	0.00
52 CM	Toluene	0.780	0.752	3.6#	77	0.00
53 T	t-1,3-Dichloropropene	0.352	0.340	3.4	78	0.00
54 T	cis-1,3-Dichloropropene	0.430	0.420	2.3	80	0.00
55 T	1,1,2-Trichloroethane	0.209	0.203	2.9	80	0.00
56 T	Ethyl methacrylate	0.284	0.282	0.7	83	0.00
57 T	1,3-Dichloropropane	0.340	0.331	2.6	80	0.00
58 T	2-Chloroethyl Vinyl ether	0.136	0.105	22.8	75	0.00
59 T	2-Hexanone	0.114	0.121	-6.1	88	0.00
60 T	Dibromochloromethane	0.271	0.261	3.7	79	0.00
61 T	1,2-Dibromoethane	0.197	0.188	4.6	80	0.00
62 S	4-Bromofluorobenzene	0.334	0.324	3.0	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.333	0.319	4.2	76	0.00
65 PM	Chlorobenzene	1.018	0.968	4.9	77	0.00
66 T	1,1,1,2-Tetrachloroethane	0.329	0.313	4.9	78	0.00
67 C	Ethyl Benzene	1.719	1.676	2.5#	79	0.00
68 T	m/p-Xylenes	0.680	0.651	4.3	77	0.00
69 T	o-Xylene	0.653	0.621	4.9	78	0.00
70 T	Styrene	1.090	1.048	3.9	78	0.00
71 P	Bromoform	0.182	0.172	5.5	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	3.749	3.629	3.2	78	0.00
74 T	N-amyl acetate	0.778	0.795	-2.2	89	0.00
75 P	1,1,2,2-Tetrachloroethane	0.598	0.570	4.7	82	0.00
76 T	1,2,3-Trichloropropane	0.400	0.412	-3.0	90	0.00
77 T	Bromobenzene	0.861	0.828	3.8	78	0.00
78 T	n-propylbenzene	4.360	4.253	2.5	78	0.00
79 T	2-Chlorotoluene	2.423	2.282	5.8	77	0.00
80 T	1,3,5-Trimethylbenzene	2.956	2.854	3.5	78	0.00
81 T	trans-1,4-Dichloro-2-butene	0.208	0.194	6.7	77	0.00
82 T	4-Chlorotoluene	2.434	2.336	4.0	79	0.00
83 T	tert-Butylbenzene	2.727	2.687	1.5	79	0.00
84 T	1,2,4-Trimethylbenzene	2.896	2.813	2.9	78	0.00
85 T	sec-Butylbenzene	3.999	3.940	1.5	79	0.00
86 T	p-Isopropyltoluene	3.290	3.237	1.6	78	0.00
87 T	1,3-Dichlorobenzene	1.663	1.602	3.7	79	0.00
88 T	1,4-Dichlorobenzene	1.621	1.557	3.9	78	0.00
89 T	n-Butylbenzene	2.978	2.979	-0.0	80	0.00
90 T	Hexachloroethane	0.645	0.619	4.0	77	0.00
91 T	1,2-Dichlorobenzene	1.430	1.424	0.4	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.081	0.079	2.5	88	0.00
93 T	1,2,4-Trichlorobenzene	0.804	0.826	-2.7	84	0.00
94 T	Hexachlorobutadiene	0.443	0.458	-3.4	80	0.00
95 T	Naphthalene	1.473	1.526	-3.6	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051024\
Data File : VY018225.D
Acq On : 10 May 2024 11:37
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: May 11 00:47:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.660	0.695	-5.3	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6