

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051123\
 Data File : VY013711.D
 Acq On : 11 May 2023 12:40
 Operator : KP/MD
 Sample : 02719-08
 Misc : 6.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-02_5-7_20230509

Quant Time: May 12 04:41:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	140065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	229387	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	222392	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	99445	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	95606	62.267	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.540%
35) Dibromofluoromethane	7.716	113	81396	56.418	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.840%
50) Toluene-d8	10.179	98	283018	55.143	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.280%
62) 4-Bromofluorobenzene	12.483	95	87126	50.354	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	100.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051123\
Data File : VY013711.D
Acq On : 11 May 2023 12:40
Operator : KP/MD
Sample : 02719-08
Misc : 6.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-02_5-7_20230509

Quant Time: May 12 04:41:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

