

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051123\
 Data File : VY013715.D
 Acq On : 11 May 2023 14:37
 Operator : KP/MD
 Sample : 02743-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DRUM-A-SLUDGE

Quant Time: May 12 04:42:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	61002	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	92001	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	88660	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	37490	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	42160	63.046	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.100%	
35) Dibromofluoromethane	7.716	113	34042	58.831	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 117.660%	
50) Toluene-d8	10.179	98	114327	55.540	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 111.080%	
62) 4-Bromofluorobenzene	12.483	95	33728	48.602	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 97.200%	
Target Compounds						
16) Acetone	3.936	43	3207	12.178	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051123\
Data File : VY013715.D
Acq On : 11 May 2023 14:37
Operator : KP/MD
Sample : 02743-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DRUM-A-SLUDGE

Quant Time: May 12 04:42:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

