

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051123\
 Data File : VY013717.D
 Acq On : 11 May 2023 15:48
 Operator : KP/MD
 Sample : 02743-05
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-COMP

Quant Time: May 12 04:43:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	196151	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	306549	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	241741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	73291	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	117995	54.875	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.760%	
35) Dibromofluoromethane	7.716	113	100899	52.332	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.660%	
50) Toluene-d8	10.179	98	368639	53.746	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 107.500%	
62) 4-Bromofluorobenzene	12.483	95	81311	35.165	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 70.320%	
Target Compounds						
16) Acetone	3.954	43	7306	8.628	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051123\
Data File : VY013717.D
Acq On : 11 May 2023 15:48
Operator : KP/MD
Sample : 02743-05
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-COMP

Quant Time: May 12 04:43:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

