



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051123\
Data File : VY013703.D
Acq On : 11 May 2023 08:49
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Title : SW846 8260
Last Update : Tue Apr 25 02:39:58 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	19850	PASS
75	95	30	60	53.3	52072	PASS
95	95	100	100	100.0	97747	PASS
96	95	5	9	6.2	6025	PASS
173	174	0.00	2	1.8	1666	PASS
174	95	50	100	94.4	92285	PASS
175	174	5	9	8.2	7551	PASS
176	174	95	101	96.0	88587	PASS
177	176	5	9	6.9	6139	PASS