

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051221\
 Data File : VY004753.D
 Acq On : 12 May 2021 13:09
 Operator : SY/MD
 Sample : M2361-01
 Misc : 6.67G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CL-02-051121

Quant Time: May 13 00:56:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051021S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 11 04:13:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	146237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	250689	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	230155	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	95699	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	95480	54.85	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.70%
35) Dibromofluoromethane	7.734	113	78333	57.08	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.16%
50) Toluene-d8	10.185	98	311834	50.30	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.60%
62) 4-Bromofluorobenzene	12.489	95	101689	48.60	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051221\
Data File : VY004753.D
Acq On : 12 May 2021 13:09
Operator : SY/MD
Sample : M2361-01
Misc : 6.67G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CL-02-051121

Quant Time: May 13 00:56:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051021S.M
Quant Title : SW846 8260
QLast Update : Tue May 11 04:13:44 2021
Response via : Initial Calibration

