


Dichlorodifluoromethane

Response Ratio



$$\text{Response} = 2.974\text{e-}001 * \text{Amt} + 4.936\text{e-}002$$


Coef of Det (r^2) = 0.995692 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y051222S.M

Calibration Table Last Updated: Fri May 13 02:08:33 2022

Cyclohexane

Response Ratio



$$\text{Response} = 7.286\text{e-}001 * \text{Amt} + 7.221\text{e-}002$$


Coef of Det (r^2) = 0.997492 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y051222S.M

Calibration Table Last Updated: Fri May 13 02:08:33 2022

Methylene Chloride

Response Ratio



$$\text{Response} = 4.428\text{e-}001 * \text{Amt} + 1.213\text{e-}001$$


Coef of Det (r^2) = 0.995285 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y051222S.M

Calibration Table Last Updated: Fri May 13 02:08:33 2022

Acetone


Response Ratio



$R = -2.412e-003 A^2 + 1.160e-001 A + 6.334e-003$
 Coef of Det (r^2) = 0.998228 Curve Fit: Quadratic
 Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y051222S.M
 Calibration Table Last Updated: Fri May 13 02:08:33 2022

Methyl Iodide

Response Ratio



$$\text{Response} = 5.043\text{e-}001 * \text{Amt} - 4.537\text{e-}003$$


Coef of Det (r^2) = 0.996296 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y051222S.M

Calibration Table Last Updated: Fri May 13 02:08:33 2022

Chloroethane


Response Ratio



$R = -8.053e-002 A^2 + 7.739e-001 A + 1.409e-002$
 Coef of Det (r^2) = 0.999204 Curve Fit: Quadratic
 Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y051222S.M
 Calibration Table Last Updated: Fri May 13 02:08:33 2022

Bromomethane


Response Ratio



$R = -1.019e-001 A^2 + 9.139e-001 A + 8.228e-002$
 Coef of Det (r^2) = 0.998784 Curve Fit: Quadratic
 Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y051222S.M
 Calibration Table Last Updated: Fri May 13 02:08:33 2022

Vinyl Chloride


Response Ratio



$R = -1.189e-001 A^2 + 1.099e+000 A + 3.203e-002$
 Coef of Det (r^2) = 0.999402 Curve Fit: Quadratic
 Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y051222S.M
 Calibration Table Last Updated: Fri May 13 02:08:33 2022


Chloromethane

Response Ratio



$R = -9.163e-002 A^2 + 7.918e-001 A + 3.360e-002$
Coef of Det (r^2) = 0.999226 Curve Fit: Quadratic
Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y051222S.M
Calibration Table Last Updated: Fri May 13 02:08:33 2022

Methyl Acetate



Response = 2.484e-001 * Amt + 5.290e-002
Coef of Det (r^2) = 0.993595 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y051222S.M
Calibration Table Last Updated: Fri May 13 02:08:33 2022