

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051324\
 Data File : VY018244.D
 Acq On : 13 May 2024 16:12
 Operator : SY/MD
 Sample : P2466-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-03-051324

Quant Time: May 14 02:29:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	165445	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	309461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	272613	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	101504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	78234	64.278	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.560%
35) Dibromofluoromethane	7.710	113	91434	56.220	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.440%
50) Toluene-d8	10.179	98	346814	54.115	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.220%
62) 4-Bromofluorobenzene	12.483	95	102848	49.692	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.380%
Target Compounds						
20) Methylene Chloride	4.710	84	21265	12.386	ug/l	Qvalue # 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051324\
Data File : VY018244.D
Acq On : 13 May 2024 16:12
Operator : SY/MD
Sample : P2466-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-051324

Quant Time: May 14 02:29:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

