

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051324\
 Data File : VY018248.D
 Acq On : 13 May 2024 17:47
 Operator : SY/MD
 Sample : P2468-03
 Misc : 5.62g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC-MH-II

Quant Time: May 14 02:31:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	175293	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	337236	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	309161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	118676	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	92841	71.993	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	143.980%
35) Dibromofluoromethane	7.710	113	100458	56.681	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.360%
50) Toluene-d8	10.179	98	383188	54.866	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.740%
62) 4-Bromofluorobenzene	12.483	95	118946	52.737	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	105.480%
Target Compounds						
20) Methylene Chloride	4.710	84	18642	10.249	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051324\
Data File : VY018248.D
Acq On : 13 May 2024 17:47
Operator : SY/MD
Sample : P2468-03
Misc : 5.62g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC-MH-II

Quant Time: May 14 02:31:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

