

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051421\
 Data File : VY004788.D
 Acq On : 14 May 2021 17:18
 Operator : SY/MD
 Sample : M2387-01
 Misc : 5.05G/5ML/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 41789

Quant Time: May 15 04:00:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051021S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 11 04:13:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	185443	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	340120	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	325113	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	138161	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	133690	60.57	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 121.14%		
35) Dibromofluoromethane	7.728	113	105236	56.52	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 113.04%		
50) Toluene-d8	10.185	98	424443	50.46	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 100.92%		
62) 4-Bromofluorobenzene	12.483	95	159968	56.35	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 112.70%		
Target Compounds						
16) Acetone	3.960	43	6289	12.68	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051421\
Data File : VY004788.D
Acq On : 14 May 2021 17:18
Operator : SY/MD
Sample : M2387-01
Misc : 5.05G/5ML/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
41789

Quant Time: May 15 04:00:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051021S.M
Quant Title : SW846 8260
QLast Update : Tue May 11 04:13:44 2021
Response via : Initial Calibration

