

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051424\
 Data File : VY018258.D
 Acq On : 14 May 2024 15:11
 Operator : SY/MD
 Sample : P2477-01
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ETGI-339

Quant Time: May 15 04:15:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	163396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	311309	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	275836	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	101155	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	82116	68.313	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.620%
35) Dibromofluoromethane	7.716	113	92983	56.833	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.660%
50) Toluene-d8	10.179	98	350546	54.372	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.740%
62) 4-Bromofluorobenzene	12.483	95	103823	49.866	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.740%
Target Compounds						
20) Methylene Chloride	4.704	84	15132	8.925	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051424\
Data File : VY018258.D
Acq On : 14 May 2024 15:11
Operator : SY/MD
Sample : P2477-01
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ETGI-339

Quant Time: May 15 04:15:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

