

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051424\
 Data File : VY018259.D
 Acq On : 14 May 2024 15:35
 Operator : SY/MD
 Sample : P2476-03
 Misc : 5.22g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1-COMP

Quant Time: May 15 04:15:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	31090	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	50927	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	36802	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	11107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	10956	47.901	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.800%
35) Dibromofluoromethane	7.716	113	14049	52.491	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.980%
50) Toluene-d8	10.179	98	55609	52.725	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	105.460%
62) 4-Bromofluorobenzene	12.483	95	12544	36.829	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	73.660%
Target Compounds						
20) Methylene Chloride	4.704	84	2150	6.664	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051424\
Data File : VY018259.D
Acq On : 14 May 2024 15:35
Operator : SY/MD
Sample : P2476-03
Misc : 5.22g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1-COMP

Quant Time: May 15 04:15:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

