

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051623\
 Data File : VY013759.D
 Acq On : 16 May 2023 15:05
 Operator : KP/MD
 Sample : 02774-04
 Misc : 5.36g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-N-1

Quant Time: May 17 07:47:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	182503	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	327482	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	329938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	154319	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	128687	64.323	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.640%	
35) Dibromofluoromethane	7.710	113	109864	53.339	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.680%	
50) Toluene-d8	10.173	98	417526	56.983	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 113.960%	
62) 4-Bromofluorobenzene	12.477	95	152361	61.680	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 123.360%	
Target Compounds						
16) Acetone	3.936	43	910	1.155	ug/l	90
20) Methylene Chloride	4.710	84	28394	4.073	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051623\
Data File : VY013759.D
Acq On : 16 May 2023 15:05
Operator : KP/MD
Sample : 02774-04
Misc : 5.36g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-N-1

Quant Time: May 17 07:47:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

