

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY051820\
 Data File : VY002703.D
 Acq On : 18 May 2020 12:05
 Operator : SY/MD
 Sample : L2663-03RE
 Misc : 6.24G/5ML/MSVOA Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11935RE

Quant Time: May 19 00:56:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y051520S.M
 Quant Title : SW846 8260
 QLast Update : Sat May 16 04:52:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	266176	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	409183	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	321834	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	109734	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.15	65	118046	56.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.40%	
35) Dibromofluoromethane	7.72	113	119919	53.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.20%	
50) Toluene-d8	10.18	98	461683	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
62) 4-Bromofluorobenzene	12.48	95	116395	37.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.50%	
Target Compounds						
3) Chloromethane	2.12	50	2963	1.356	ug/l	98
11) Tert butyl alcohol	4.96	59	8471	40.062	ug/l #	94
68) m/p-Xylenes	11.70	106	8795	2.070	ug/l	98
69) o-Xylene	12.03	106	16031	4.021	ug/l	100
80) 1,3,5-Trimethylbenzene	12.81	105	78730	13.068	ug/l	97
84) 1,2,4-Trimethylbenzene	13.12	105	68215	11.329	ug/l	99
86) p-Isopropyltoluene	13.37	119	20923	3.038	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY051820\
Data File : VY002703.D
Acq On : 18 May 2020 12:05
Operator : SY/MD
Sample : L2663-03RE
Misc : 6.24G/5ML/MSVOA Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11935RE

Quant Time: May 19 00:56:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y051520S.M
Quant Title : SW846 8260
QLast Update : Sat May 16 04:52:36 2020
Response via : Initial Calibration

