

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051821\
 Data File : VY004821.D
 Acq On : 18 May 2021 12:45
 Operator : SY/MD
 Sample : M2392-10
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-04-4.5-5.0

Quant Time: May 19 01:02:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051021S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 11 04:13:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	91948	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	160165	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	145974	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	62748	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	55831	51.01	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.02%
35) Dibromofluoromethane	7.728	113	49177	56.09	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.18%
50) Toluene-d8	10.185	98	197438	49.85	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.70%
62) 4-Bromofluorobenzene	12.483	95	60438	45.21	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.42%
Target Compounds						
16) Acetone	3.960	43	2340	9.52	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051821\
Data File : VY004821.D
Acq On : 18 May 2021 12:45
Operator : SY/MD
Sample : M2392-10
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampled :
SB-04-4.5-5.0

Quant Time: May 19 01:02:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051021S.M
Quant Title : SW846 8260
QLast Update : Tue May 11 04:13:44 2021
Response via : Initial Calibration

