

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051822\
 Data File : VY008852.D
 Acq On : 19 May 2022 06:53
 Operator : KP/MD
 Sample : N2841-20
 Misc : 4.69g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P009-SS003-2430-01

Quant Time: May 19 08:10:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 07:32:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	119306	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	213379	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	204859	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	88269	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	78328	70.769	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	141.540%
35) Dibromofluoromethane	7.722	113	67962	54.512	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.020%
50) Toluene-d8	10.179	98	267921	48.451	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.900%
62) 4-Bromofluorobenzene	12.483	95	78249	40.294	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.580%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051822\
Data File : VY008852.D
Acq On : 19 May 2022 06:53
Operator : KP/MD
Sample : N2841-20
Misc : 4.69g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P009-SS003-2430-01

Quant Time: May 19 08:10:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 07:32:50 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008852.D
Acq: 19 May 2022 06:53




Instrument :
MSVOA_Y
ClientSampleId :
P009-SS003-2430-01


Tgt Ion:168 Resp: 119306
Ion Ratio Lower Upper
168 100
99 54.2 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 70.769 ug/l
RT: 8.142 min Scan# 1037
Delta R.T. -0.000 min
Lab File: VY008852.D
Acq: 19 May 2022 06:53

Tgt Ion: 65 Resp: 78328
Ion Ratio Lower Upper
65 100
67 52.2 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. -0.000 min

Lab File: VY008852.D

Acq: 19 May 2022 06:53

Instrument :

MSVOA_Y

ClientSampleId :

P009-SS003-2430-01





Tgt Ion:114 Resp: 213379

Ion Ratio Lower Upper

114 100

63 18.5 0.0 43.2

88 13.7 0.0 31.8



#35

Dibromofluoromethane

Concen: 54.512 ug/l

RT: 7.722 min Scan# 968

Delta R.T. 0.006 min

Lab File: VY008852.D

Acq: 19 May 2022 06:53




Tgt Ion:113 Resp: 67962


Ion Ratio Lower Upper

113 100

111 104.6 82.6 124.0

192 19.6 17.6 26.4








#50
Toluene-d8
Concen: 48.451 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.001 min
Lab File: VY008852.D
Acq: 19 May 2022 06:53




Instrument :
MSVOA_Y
ClientSampleId :
P009-SS003-2430-01

Tgt Ion: 98 Resp: 267921
Ion Ratio Lower Upper
98 100
100 64.7 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 40.294 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY008852.D
Acq: 19 May 2022 06:53

Tgt Ion: 95 Resp: 78249
Ion Ratio Lower Upper
95 100
174 91.9 0.0 177.0
176 86.9 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 117
Delta R.T. -0.000 min
Lab File: VY008852.D
Acq: 19 May 2022 06:53

Instrument :
MSVOA_Y
ClientSampleId :
P009-SS003-2430-01

Tgt Ion:	117	Resp:	204859
Ion	Ratio	Lower	Upper
117	100		
82	52.2	43.0	64.4
119	32.1	25.4	38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY008852.D
Acq: 19 May 2022 06:53

Tgt Ion:	152	Resp:	88269
Ion	Ratio	Lower	Upper
152	100		
115	55.7	28.2	84.7
150	158.8	0.0	347.6

