

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051822\
 Data File : VY008855.D
 Acq On : 19 May 2022 08:01
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
 Supervised By :Mahesh Dadoda 05/23/2022

Quant Time: May 19 09:16:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 07:32:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	138523	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	244280	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	251743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	148811	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	74790	58.198	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.400%
35) Dibromofluoromethane	7.716	113	69777	48.888	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.780%
50) Toluene-d8	10.179	98	293908	46.560	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.120%
62) 4-Bromofluorobenzene	12.483	95	108661	48.135	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.260%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.906	85	40350	46.381	ug/l	98
3) Chloromethane	2.113	50	66469	53.884	ug/l	96
4) Vinyl Chloride	2.253	62	91467	48.446	ug/l	97
5) Bromomethane	2.650	94	86232	53.466	ug/l	99
6) Chloroethane	2.790	64	68267	47.265	ug/l	99
7) Trichlorofluoromethane	3.125	101	85637	45.647	ug/l	90
8) Diethyl Ether	3.528	74	31850	51.893	ug/l	80
9) 1,1,2-Trichlorotrifluo...	3.899	101	56019	48.995	ug/l	94
10) Methyl Iodide	4.088	142	62055	43.240	ug/l	92
11) Tert butyl alcohol	4.954	59	27938	196.707	ug/l #	92
12) 1,1-Dichloroethene	3.875	96	51499	43.814	ug/l	79
13) Acrolein	3.723	56	7591	108.040	ug/l	97
14) Allyl chloride	4.479	41	82715	47.841	ug/l #	88
15) Acrylonitrile	5.161	53	91003	218.651	ug/l	98
16) Acetone	3.948	43	62907	215.750	ug/l	95
17) Carbon Disulfide	4.198	76	142171	46.878	ug/l	99
18) Methyl Acetate	4.473	43	45402	49.715	ug/l #	90
19) Methyl tert-butyl Ether	5.222	73	177319	56.013	ug/l	95
20) Methylene Chloride	4.716	84	75069	57.685	ug/l	87
21) trans-1,2-Dichloroethene	5.216	96	67837	47.069	ug/l	86
22) Diisopropyl ether	6.119	45	217659	59.445	ug/l #	96
23) Vinyl Acetate	6.064	43	697464	270.162	ug/l #	92
24) 1,1-Dichloroethane	6.015	63	128158	52.749	ug/l	99
25) 2-Butanone	6.984	43	118638	219.985	ug/l #	86
26) 2,2-Dichloropropane	6.978	77	95711	41.448	ug/l	93
27) cis-1,2-Dichloroethene	6.984	96	83826	52.440	ug/l	86
28) Bromochloromethane	7.332	49	51746	56.178	ug/l #	87
29) Tetrahydrofuran	7.350	42	80020	222.737	ug/l #	86
30) Chloroform	7.509	83	141000	54.129	ug/l	97
31) Cyclohexane	7.789	56	99986	49.207	ug/l	91
32) 1,1,1-Trichloroethane	7.704	97	119800	49.604	ug/l	98
36) 1,1-Dichloropropene	7.917	75	98199	42.957	ug/l	98
37) Ethyl Acetate	7.070	43	55256	40.563	ug/l	97
38) Carbon Tetrachloride	7.899	117	105247	42.960	ug/l	98
39) Methylcyclohexane	9.185	83	119496	44.079	ug/l	89
40) Benzene	8.161	78	308514	48.233	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051822\
 Data File : VY008855.D
 Acq On : 19 May 2022 08:01
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/23/2022
 Supervised By : Mahesh Dadoda 05/23/2022

Quant Time: May 19 09:16:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 07:32:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.320	41	28663m	40.010	ug/l	92
42) 1,2-Dichloroethane	8.240	62	93051	49.559	ug/l	92
43) Isopropyl Acetate	8.271	43	107413	47.366	ug/l #	90
44) Trichloroethene	8.941	130	79568	45.730	ug/l	95
45) 1,2-Dichloropropane	9.216	63	78851	51.871	ug/l	97
46) Dibromomethane	9.307	93	48584	48.638	ug/l	96
47) Bromodichloromethane	9.496	83	115490	52.153	ug/l	98
48) Methyl methacrylate	9.289	41	48036	48.241	ug/l #	83
49) 1,4-Dioxane	9.295	88	12676	844.854	ug/l #	78
51) 4-Methyl-2-Pentanone	10.069	43	306621	229.035	ug/l	91
52) Toluene	10.246	92	221564	50.976	ug/l	96
53) t-1,3-Dichloropropene	10.465	75	117792	51.211	ug/l	98
54) cis-1,3-Dichloropropene	9.929	75	124231	50.086	ug/l	90
55) 1,1,2-Trichloroethane	10.642	97	71448	52.599	ug/l	97
56) Ethyl methacrylate	10.508	69	91066	55.258	ug/l #	83
57) 1,3-Dichloropropane	10.788	76	115385	51.661	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.783	63	227449	270.650	ug/l	96
59) 2-Hexanone	10.831	43	213207	222.892	ug/l	90
60) Dibromochloromethane	10.984	129	87087	53.406	ug/l	99
61) 1,2-Dibromoethane	11.087	107	69617	51.393	ug/l	97
64) Tetrachloroethene	10.721	164	69829	43.333	ug/l	96
65) Chlorobenzene	11.514	112	243074	47.709	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.593	131	94405	50.718	ug/l	97
67) Ethyl Benzene	11.593	91	436641	47.989	ug/l	99
68) m/p-Xylenes	11.703	106	358350	98.860	ug/l	92
69) o-Xylene	12.032	106	171503	51.206	ug/l	90
70) Styrene	12.044	104	302202	53.301	ug/l	94
71) Bromoform	12.209	173	61982	49.052	ug/l #	100
73) Isopropylbenzene	12.331	105	459369	45.010	ug/l	100
74) N-amyl acetate	12.142	43	109716	45.038	ug/l #	85
75) 1,1,2,2-Tetrachloroethane	12.581	83	98050	43.214	ug/l	99
76) 1,2,3-Trichloropropane	12.636	75	73415m	42.709	ug/l	
77) Bromobenzene	12.611	156	108236	45.769	ug/l	93
78) n-propylbenzene	12.672	91	559505	44.624	ug/l	99
79) 2-Chlorotoluene	12.758	91	317652	46.775	ug/l	96
80) 1,3,5-Trimethylbenzene	12.813	105	399577	48.490	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.380	75	29426	37.443	ug/l #	85
82) 4-Chlorotoluene	12.855	91	340683	47.486	ug/l	96
83) tert-Butylbenzene	13.075	119	354486	47.834	ug/l	94
84) 1,2,4-Trimethylbenzene	13.123	105	404762	49.617	ug/l	98
85) sec-Butylbenzene	13.258	105	549200	47.500	ug/l	99
86) p-Isopropyltoluene	13.373	119	453043	48.055	ug/l	98
87) 1,3-Dichlorobenzene	13.367	146	237558	47.809	ug/l	99
88) 1,4-Dichlorobenzene	13.447	146	236902	48.200	ug/l	98
89) n-Butylbenzene	13.697	91	397197	44.386	ug/l	99
90) Hexachloroethane	13.965	117	86802	48.698	ug/l	84
91) 1,2-Dichlorobenzene	13.739	146	213338	49.096	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.355	75	15001	38.667	ug/l	82
93) 1,2,4-Trichlorobenzene	15.007	180	108706	45.812	ug/l	100
94) Hexachlorobutadiene	15.117	225	59325	43.448	ug/l	99
95) Naphthalene	15.239	128	236646	44.611	ug/l	99
96) 1,2,3-Trichlorobenzene	15.428	180	98440	47.225	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051822\
 Data File : VY008855.D
 Acq On : 19 May 2022 08:01
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
 Supervised By :Mahesh Dadoda 05/23/2022


Quant Time: May 19 09:16:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 07:32:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


05/23/2022
 Supervised By :Mahesh
 Dadoda

05/23/2022



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 13852
Ion Ratio Lower Upper
168 100
99 50.9 46.9 70.3


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#2
Dichlorodifluoromethane
Concen: 46.381 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 85 Resp: 40350
Ion Ratio Lower Upper
85 100
87 33.2 16.1 48.2





#3
Chloromethane
Concen: 53.884 ug/l
RT: 2.113 min Scan# 41
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 50 Resp: 66469
Ion Ratio Lower Upper
50 100
52 33.2 24.6 37.0


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#4
Vinyl Chloride
Concen: 48.446 ug/l
RT: 2.253 min Scan# 71
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Tgt Ion: 62 Resp: 91467
Ion Ratio Lower Upper
62 100
64 29.8 25.0 37.6







#5
Bromomethane
Concen: 53.466 ug/l
RT: 2.650 min Scan# 11
Delta R.T. 0.007 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 94 Resp: 8623
Ion Ratio Lower Upper
94 100
96 93.2 73.5 110.3



Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#6
Chloroethane
Concen: 47.265 ug/l
RT: 2.790 min Scan# 159
Delta R.T. -0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Tgt Ion: 64 Resp: 68267
Ion Ratio Lower Upper
64 100
66 31.2 25.4 38.0







#7
Trichlorofluoromethane
Concen: 45.647 ug/l
RT: 3.125 min Scan# 213
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 8563
Ion Ratio Lower Upper
101 100
103 67.9 48.0 72.0



Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#8
Diethyl Ether
Concen: 51.893 ug/l
RT: 3.528 min Scan# 280
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Tgt Ion: 74 Resp: 31850
Ion Ratio Lower Upper
74 100
45 89.3 55.1 165.5







#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.995 ug/l
RT: 3.899 min Scan# 340
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 56019
Ion Ratio Lower Upper
101 100
85 46.9 40.6 60.8
151 75.4 64.9 97.3



Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#10
Methyl Iodide
Concen: 43.240 ug/l
RT: 4.088 min Scan# 372
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion:142 Resp: 62055
Ion Ratio Lower Upper
142 100
127 44.7 41.5 62.3
141 14.9 11.9 17.9





#11
Tert butyl alcohol
Concen: 196.707 ug/l
RT: 4.954 min Scan# 512
Delta R.T. 0.012 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 59 Resp: 27938
Ion Ratio Lower Upper
59 100
57 10.8 6.4 9.6


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#12
1,1-Dichloroethene
Concen: 43.814 ug/l
RT: 3.875 min Scan# 337
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 96 Resp: 51499
Ion Ratio Lower Upper
96 100
61 146.4 145.8 218.6
98 59.1 52.5 78.7





#13
Acrolein
Concen: 108.040 ug/l
RT: 3.723 min Scan# 313
Delta R.T. -0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 759
Ion Ratio Lower Upper
56 100
55 74.4 57.6 86.4


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#14
Allyl chloride
Concen: 47.841 ug/l
RT: 4.479 min Scan# 436
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 41 Resp: 82715
Ion Ratio Lower Upper
41 100
39 80.6 58.8 88.2
76 42.7 25.1 37.7#





#15
Acrylonitrile
Concen: 218.651 ug/l
RT: 5.161 min Scan# 547
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 53 Resp: 9100
Ion Ratio Lower Upper
53 100
52 81.1 67.0 100.4
51 36.5 29.8 44.6


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#16
Acetone
Concen: 215.750 ug/l
RT: 3.948 min Scan# 349
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 43 Resp: 62907
Ion Ratio Lower Upper
43 100
58 30.1 22.0 33.0





#17
Carbon Disulfide
Concen: 46.878 ug/l
RT: 4.198 min Scan# 389
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 14217
Ion Ratio Lower Upper
76 100
78 8.8 7.4 11.0


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#18
Methyl Acetate
Concen: 49.715 ug/l
RT: 4.473 min Scan# 435
Delta R.T. 0.001 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 43 Resp: 45402
Ion Ratio Lower Upper
43 100
74 25.3 16.6 24.8





#19
Methyl tert-butyl Ether
Concen: 56.013 ug/l
RT: 5.222 min Scan# 51
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 73 Resp: 177319
Ion Ratio Lower Upper
73 100
57 20.9 18.6 27.8


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#20
Methylene Chloride
Concen: 57.685 ug/l
RT: 4.716 min Scan# 475
Delta R.T. 0.012 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 84 Resp: 75069
Ion Ratio Lower Upper
84 100
49 120.9 114.4 171.6
51 36.2 35.3 52.9
86 65.4 53.7 80.5





#21
trans-1,2-Dichloroethene
Concen: 47.069 ug/l
RT: 5.216 min Scan# 51
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 6783
Ion Ratio Lower Upper
96 100
61 127.8 120.5 180.7
98 65.2 49.8 74.8


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#22
Diisopropyl ether
Concen: 59.445 ug/l
RT: 6.119 min Scan# 705
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 45 Resp: 217659
Ion Ratio Lower Upper
45 100
43 61.0 49.3 73.9
87 29.0 19.2 28.8#
59 12.5 8.4 12.6





#23
 Vinyl Acetate
 Concen: 270.162 ug/l
 RT: 6.064 min Scan# 695
 Delta R.T. 0.006 min
 Lab File: VY008855.D
 Acq: 19 May 2022 08:01



Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion: 43 Resp: 697464
 Ion Ratio Lower Upper
 43 100
 86 11.6 7.1 10.7


Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
 Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
 Supervised By :Mahesh
 Dadoda


05/23/2022



#24
 1,1-Dichloroethane
 Concen: 52.749 ug/l
 RT: 6.015 min Scan# 688
 Delta R.T. 0.000 min
 Lab File: VY008855.D
 Acq: 19 May 2022 08:01


Tgt Ion: 63 Resp: 128158
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.6 10.8
 100 4.5 2.0 5.8





#25
2-Butanone
Concen: 219.985 ug/l
RT: 6.984 min Scan# 847
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 118638
Ion Ratio Lower Upper
43 100
72 28.9 17.7 26.5


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#26
2,2-Dichloropropane
Concen: 41.448 ug/l
RT: 6.978 min Scan# 846
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01




Tgt Ion: 77 Resp: 95711
Ion Ratio Lower Upper
77 100
97 24.1 10.4 31.4





#27
cis-1,2-Dichloroethene
Concen: 52.440 ug/l
RT: 6.984 min Scan# 847
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 83820
Ion Ratio Lower Upper
96 100
61 135.8 0.0 319.8
98 65.3 0.0 127.0


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#28
Bromochloromethane
Concen: 56.178 ug/l
RT: 7.332 min Scan# 904
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 49 Resp: 51746
Ion Ratio Lower Upper
49 100
129 2.2 0.0 2.2#
130 79.8 55.4 83.0





#29
Tetrahydrofuran
Concen: 222.737 ug/l
RT: 7.350 min Scan# 907
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
Client Sample Id :
VSTDCCC050EC



Tgt Ion: 42 Resp: 80020
Ion Ratio Lower Upper
42 100
72 46.7 30.3 45.5
71 44.7 29.0 43.4


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#30
Chloroform
Concen: 54.129 ug/l
RT: 7.509 min Scan# 933
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01




Tgt Ion: 83 Resp: 141000
Ion Ratio Lower Upper
83 100
85 62.7 51.8 77.6





#31
Cyclohexane
Concen: 49.207 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 9998
Ion Ratio Lower Upper
56 100
69 34.3 26.7 40.1
84 95.3 67.2 100.8


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#32
1,1,1-Trichloroethane
Concen: 49.604 ug/l
RT: 7.704 min Scan# 965
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Tgt Ion: 97 Resp: 119800
Ion Ratio Lower Upper
97 100
99 63.8 50.7 76.1
61 43.9 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 58.198 ug/l

RT: 8.137 min Scan# 1036

Delta R.T. -0.006 min

Lab File: VY008855.D

Acq: 19 May 2022 08:01

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 74790

Ion Ratio Lower Upper

65 100



67 52.6 0.0 99.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022



Supervised By :Mahesh Dadoda 05/23/2022



05/23/2022

Supervised By :Mahesh
Dadoda

05/23/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY008855.D

Acq: 19 May 2022 08:01




Tgt Ion:114 Resp: 244280


Ion Ratio Lower Upper

114 100

63 18.2 0.0 43.2


88 13.3 0.0 31.8





#35
Dibromofluoromethane
Concen: 48.888 ug/l
RT: 7.716 min Scan# 90
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 113 Resp: 6977
Ion Ratio Lower Upper
113 100
111 104.7 82.6 124.0
192 19.2 17.6 26.4


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#36
1,1-Dichloropropene
Concen: 42.957 ug/l
RT: 7.917 min Scan# 1000
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01




Tgt Ion: 75 Resp: 98199
Ion Ratio Lower Upper
75 100
110 35.7 17.1 51.2
77 30.6 24.4 36.6





#37
Ethyl Acetate
Concen: 40.563 ug/l
RT: 7.070 min Scan# 80
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 55250
Ion Ratio Lower Upper
43 100
61 13.7 9.7 14.5
70 9.6 7.1 10.7


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#38
Carbon Tetrachloride
Concen: 42.960 ug/l
RT: 7.899 min Scan# 997
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion:117 Resp: 105247
Ion Ratio Lower Upper
117 100
119 99.3 77.9 116.9
121 30.8 24.8 37.2





#39
Methylcyclohexane
Concen: 44.079 ug/l
RT: 9.185 min Scan# 119490
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 119490
Ion Ratio Lower Upper
83 100
55 69.3 66.5 99.7
98 49.2 37.9 56.9


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#40
Benzene
Concen: 48.233 ug/l
RT: 8.161 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 78 Resp: 308514
Ion Ratio Lower Upper
78 100
77 23.3 20.4 30.6





#41
Methacrylonitrile
Concen: 40.010 ug/l m
RT: 7.320 min Scan# 902
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 41 Resp: 2866
Ion Ratio Lower Upper
41 100
39 82.5 92.9 139.3
67 0.0 0.0 0.0
52 0.0 0.0 0.0


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#42
1,2-Dichloroethane
Concen: 49.559 ug/l
RT: 8.240 min Scan# 1053
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 62 Resp: 93051
Ion Ratio Lower Upper
62 100
98 10.1 0.0 14.8





#43
Isopropyl Acetate
Concen: 47.366 ug/l
RT: 8.271 min Scan# 1058
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 10741
Ion Ratio Lower Upper
43 100
61 30.3 20.2 30.2
87 14.7 8.9 13.3


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#44
Trichloroethene
Concen: 45.730 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Tgt Ion:130 Resp: 79568
Ion Ratio Lower Upper
130 100
95 98.5 0.0 188.2





#45

1,2-Dichloropropane

Concen: 51.871 ug/l

RT: 9.216 min Scan# 1213

Delta R.T. 0.000 min

Lab File: VY008855.D

Acq: 19 May 2022 08:01

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 7885

Ion Ratio Lower Upper

63 100



65 30.6 23.0 34.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022



Supervised By :Mahesh Dadoda 05/23/2022



05/23/2022

Supervised By :Mahesh
Dadoda

05/23/2022



#46

Dibromomethane

Concen: 48.638 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VY008855.D

Acq: 19 May 2022 08:01




Tgt Ion: 93 Resp: 48584


Ion Ratio Lower Upper

93 100

95 81.9 64.3 96.5


174 98.3 83.4 125.0





#47
Bromodichloromethane
Concen: 52.153 ug/l
RT: 9.496 min Scan# 1259
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 115490
Ion Ratio Lower Upper
83 100
85 66.5 51.9 77.9
127 7.8 6.6 10.0


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#48
Methyl methacrylate
Concen: 48.241 ug/l
RT: 9.289 min Scan# 1225
Delta R.T. -0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 41 Resp: 48036
Ion Ratio Lower Upper
41 100
69 93.2 58.2 87.2#
39 64.6 47.4 71.0





#49
1,4-Dioxane
Concen: 844.854 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 88 Resp: 12670
Ion Ratio Lower Upper
88 100
43 25.8 32.8 49.2
58 56.9 59.8 89.6


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#50
Toluene-d8
Concen: 46.560 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Tgt Ion: 98 Resp: 293908
Ion Ratio Lower Upper
98 100
100 65.2 50.2 75.2







#51
4-Methyl-2-Pentanone
Concen: 229.035 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 30662
Ion Ratio Lower Upper
43 100
58 40.5 28.1 42.1



Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#52
Toluene
Concen: 50.976 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 92 Resp: 221564
Ion Ratio Lower Upper
92 100
91 172.6 142.6 213.8





#53
t-1,3-Dichloropropene
Concen: 51.211 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 117792
Ion Ratio Lower Upper
75 100
77 30.7 25.4 38.2


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#54
cis-1,3-Dichloropropene
Concen: 50.086 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 75 Resp: 124231
Ion Ratio Lower Upper
75 100
77 33.3 25.6 38.4
39 48.4 47.6 71.4





#55
1,1,2-Trichloroethane
Concen: 52.599 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion: 97 Resp: 71448

Ion	Ratio	Lower	Upper
97	100		
83	81.8	69.8	104.8
85	53.6	44.3	66.5
99	62.3	49.8	74.6


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda

05/23/2022






#56
Ethyl methacrylate
Concen: 55.258 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Tgt Ion: 69 Resp: 91066


Ion	Ratio	Lower	Upper
69	100		
41	62.0	63.4	95.0#
39	36.6	35.8	53.8





#57
1,3-Dichloropropane
Concen: 51.661 ug/l
RT: 10.788 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 115385
Ion Ratio Lower Upper
76 100
78 32.7 25.0 37.6


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#58
2-Chloroethyl Vinyl ether
Concen: 270.650 ug/l
RT: 9.783 min Scan# 1306
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 63 Resp: 227449
Ion Ratio Lower Upper
63 100
106 28.9 21.5 32.3





#59
2-Hexanone
Concen: 222.892 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 21320
Ion Ratio Lower Upper
43 100
58 56.0 24.4 73.4


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#60
Dibromochloromethane
Concen: 53.406 ug/l
RT: 10.984 min Scan# 1503
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Tgt Ion:129 Resp: 87087
Ion Ratio Lower Upper
129 100
127 76.8 38.7 116.1





#61
1,2-Dibromoethane
Concen: 51.393 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Instrument :

MSVOA_Y



ClientSampleId :

VSTDCCC050EC

Tgt Ion:107 Resp: 6961
Ion Ratio Lower Upper
107 100
109 91.8 75.4 113.2



Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#62
4-Bromofluorobenzene
Concen: 48.135 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 95 Resp: 108661
Ion Ratio Lower Upper
95 100
174 90.6 0.0 177.0
176 87.1 0.0 175.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 117
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:117 Resp: 25174
Ion Ratio Lower Upper
117 100
82 52.7 43.0 64.4
119 32.0 25.4 38.0


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#64
Tetrachloroethene
Concen: 43.333 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01




Tgt Ion:164 Resp: 69829
Ion Ratio Lower Upper
164 100
166 131.2 100.4 150.6
129 96.1 71.4 107.0
131 88.1 70.8 106.2





#65
Chlorobenzene
Concen: 47.709 ug/l
RT: 11.514 min Scan# 1150
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:112 Resp: 243074
Ion Ratio Lower Upper
112 100
114 31.3 25.8 38.6


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 50.718 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01




Tgt Ion:131 Resp: 94405
Ion Ratio Lower Upper
131 100
133 94.8 47.9 143.6
119 65.4 30.8 92.3





#67
Ethyl Benzene
Concen: 47.989 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 43664
Ion Ratio Lower Upper
91 100
106 31.4 24.5 36.7


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#68
m/p-Xylenes
Concen: 98.860 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Tgt Ion:106 Resp: 358350
Ion Ratio Lower Upper
106 100
91 195.5 166.6 249.8







#69
o-Xylene
Concen: 51.206 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 17150
Ion Ratio Lower Upper
106 100
91 205.1 110.3 331.0



Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#70
Styrene
Concen: 53.301 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion:104 Resp: 302202
Ion Ratio Lower Upper
104 100
78 45.2 42.4 63.6
103 54.3 44.4 66.6





#71
Bromoform
Concen: 49.052 ug/l
RT: 12.209 min Scan# 1704
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:173 Resp: 61982
Ion Ratio Lower Upper
173 100
175 49.3 24.6 74.0
254 0.0 0.0 0.0


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01




Tgt Ion:152 Resp: 148811
Ion Ratio Lower Upper
152 100
115 57.9 28.2 84.7
150 175.6 0.0 347.6





#73
Isopropylbenzene
Concen: 45.010 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 105 Resp: 459369
Ion Ratio Lower Upper
105 100
120 26.9 13.4 40.2


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#74
N-ethyl acetate
Concen: 45.038 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01




Tgt Ion: 43 Resp: 109716
Ion Ratio Lower Upper
43 100
70 47.8 28.7 43.1#
55 26.4 17.3 25.9#
61 26.4 17.0 25.6#





#75
1,1,2,2-Tetrachloroethane
Concen: 43.214 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 98050
Ion Ratio Lower Upper
83 100
131 10.0 5.6 16.8
85 64.3 32.3 96.8


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#76
1,2,3-Trichloropropane
Concen: 42.709 ug/l m
RT: 12.636 min Scan# 1774
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Tgt Ion: 75 Resp: 73415
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0







#77
Bromobenzene
Concen: 45.769 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:156 Resp: 108230
Ion Ratio Lower Upper
156 100
77 175.7 95.5 286.4
158 96.2 48.0 144.0



Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#78
n-propylbenzene
Concen: 44.624 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Tgt Ion: 91 Resp: 559505
Ion Ratio Lower Upper
91 100
120 23.2 11.3 33.9







#79
2-Chlorotoluene
Concen: 46.775 ug/l
RT: 12.758 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 317652
Ion Ratio Lower Upper
91 100
126 35.9 16.9 50.6



Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#80
1,3,5-Trimethylbenzene
Concen: 48.490 ug/l
RT: 12.813 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion:105 Resp: 399577
Ion Ratio Lower Upper
105 100
120 49.8 24.4 73.4





#81
trans-1,4-Dichloro-2-butene
Concen: 37.443 ug/l
RT: 12.380 min Scan# 1732
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 29420
Ion Ratio Lower Upper
75 100
53 85.1 85.3 127.9
89 48.0 37.0 55.4


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#82
4-Chlorotoluene
Concen: 47.486 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion: 91 Resp: 340683
Ion Ratio Lower Upper
91 100
126 36.5 17.0 51.0





#83
tert-Butylbenzene
Concen: 47.834 ug/l
RT: 13.075 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:119 Resp: 354480
Ion Ratio Lower Upper
119 100
91 63.1 34.8 104.4
134 27.7 13.3 39.9


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#84
1,2,4-Trimethylbenzene
Concen: 49.617 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Tgt Ion:105 Resp: 404762
Ion Ratio Lower Upper
105 100
120 45.7 22.3 66.9







#85
 sec-Butylbenzene
 Concen: 47.500 ug/l
 RT: 13.258 min Scan# 1875
 Delta R.T. 0.006 min
 Lab File: VY008855.D
 Acq: 19 May 2022 08:01

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion:105 Resp: 549200
 Ion Ratio Lower Upper
 105 100
 134 19.9 10.2 30.4



Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
 Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
 Supervised By :Mahesh
 Dadoda


05/23/2022



#86
 p-Isopropyltoluene
 Concen: 48.055 ug/l
 RT: 13.373 min Scan# 1895
 Delta R.T. 0.006 min
 Lab File: VY008855.D
 Acq: 19 May 2022 08:01


Tgt Ion:119 Resp: 453043
 Ion Ratio Lower Upper
 119 100
 134 26.6 13.2 39.5
 91 22.9 12.4 37.2





#87
1,3-Dichlorobenzene
Concen: 47.809 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 237558
Ion Ratio Lower Upper
146 100
111 38.1 19.5 58.5
148 62.8 31.6 95.0


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#88
1,4-Dichlorobenzene
Concen: 48.200 ug/l
RT: 13.447 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01




Tgt Ion:146 Resp: 236902
Ion Ratio Lower Upper
146 100
111 37.5 19.3 57.8
148 63.6 32.6 97.8





#89
n-Butylbenzene
Concen: 44.386 ug/l
RT: 13.697 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 397197
Ion Ratio Lower Upper
91 100
92 53.3 26.3 78.9
134 26.1 13.0 38.9


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022






05/23/2022
Supervised By :Mahesh
Dadoda


05/23/2022



#90
Hexachloroethane
Concen: 48.698 ug/l
RT: 13.965 min Scan# 1992
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01


Tgt Ion:117 Resp: 86802
Ion Ratio Lower Upper
117 100
201 71.5 43.0 128.8





#91
1,2-Dichlorobenzene
Concen: 49.096 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 213338
Ion Ratio Lower Upper
146 100
111 39.8 20.3 60.9
148 63.7 32.3 96.9


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 38.667 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01




Tgt Ion: 75 Resp: 15001
Ion Ratio Lower Upper
75 100
155 89.4 38.0 113.9
157 115.1 47.8 143.3





#93
1,2,4-Trichlorobenzene
Concen: 45.812 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:180 Resp: 108700
Ion Ratio Lower Upper
180 100
182 94.2 47.3 141.9
145 29.2 14.6 43.8


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda



05/23/2022




#94
Hexachlorobutadiene
Concen: 43.448 ug/l
RT: 15.117 min Scan# 2181
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01




Tgt Ion:225 Resp: 59325
Ion Ratio Lower Upper
225 100
223 62.7 30.8 92.3
227 63.5 31.6 94.7





#95
Naphthalene
Concen: 44.611 ug/l
RT: 15.239 min Scan# 21
Delta R.T. 0.000 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:128 Resp: 236640
Ion Ratio Lower Upper
128 100
127 13.1 11.0 16.4
129 11.0 8.4 12.6


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2022
Supervised By :Mahesh Dadoda 05/23/2022




05/23/2022
Supervised By :Mahesh
Dadoda

05/23/2022



#96
1,2,3-Trichlorobenzene
Concen: 47.225 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. 0.006 min
Lab File: VY008855.D
Acq: 19 May 2022 08:01



Tgt Ion:180 Resp: 98440
Ion Ratio Lower Upper
180 100
182 94.1 47.3 141.9
145 31.4 15.3 45.9

