

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051923\
 Data File : VY013824.D
 Acq On : 19 May 2023 16:28
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: May 22 04:21:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 18 06:54:18 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	174442	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	298969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	290299	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	133689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	119594	59.017	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 118.040%	
35) Dibromofluoromethane	7.716	113	100375	52.356	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.720%	
50) Toluene-d8	10.179	98	366385	50.409	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.820%	
62) 4-Bromofluorobenzene	12.477	95	116625	48.308	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 96.620%	
Target Compounds						
					Qvalue	
16) Acetone	3.948	43	1367	2.236	ug/l #	68
20) Methylene Chloride	4.716	84	23484	7.528	ug/l	90
44) Trichloroethene	8.947	130	3640	1.809	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051923\
Data File : VY013824.D
Acq On : 19 May 2023 16:28
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: May 22 04:21:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
Quant Title : SW846 8260
QLast Update : Thu May 18 06:54:18 2023
Response via : Initial Calibration

