

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv052020\
 Data File : VY002719.D
 Acq On : 20 May 2020 10:07
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 21 03:18:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y051520S.M
 Quant Title : SW846 8260
 QLast Update : Sat May 16 04:52:36 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	119	0.00
2 T	Dichlorodifluoromethane	0.358	0.365	-2.0	128	0.00
3 P	Chloromethane	0.410	0.405	1.2	123	0.00
4 C	Vinyl Chloride	0.428	0.445	-4.0#	127	0.00
5 T	Bromomethane	0.317	0.304	4.1	117	0.00
6 T	Chloroethane	0.273	0.274	-0.4	120	0.00
7 T	Trichlorofluoromethane	0.721	0.738	-2.4	123	0.00
8 T	Diethyl Ether	0.245	0.251	-2.4	117	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.437	0.456	-4.3	126	0.00
10 T	Methyl Iodide	0.601	0.641	-6.7	117	0.00
11 T	Tert butyl alcohol	0.040	0.036	10.0	109	0.00
12 CM	1,1-Dichloroethene	0.439	0.452	-3.0#	123	0.00
13 T	Acrolein	0.047	0.052	-10.6	131	0.00
14 T	Allyl chloride	0.630	0.662	-5.1	124	0.00
15 T	Acrylonitrile	0.111	0.110	0.9	111	0.00
16 T	Acetone	0.081	0.072	11.1	100	0.00
17 T	Carbon Disulfide	1.376	1.395	-1.4	120	0.00
18 T	Methyl Acetate	0.230	0.229	0.4	113	0.00
19 T	Methyl tert-butyl Ether	1.088	1.107	-1.7	117	0.00
20 T	Methylene Chloride	0.487	0.469	3.7	117	0.00
21 T	trans-1,2-Dichloroethene	0.488	0.494	-1.2	120	0.00
22 T	Diisopropyl ether	1.249	1.302	-4.2	121	0.00
23 T	Vinyl Acetate	0.800	0.817	-2.1	116	0.00
24 P	1,1-Dichloroethane	0.746	0.780	-4.6	122	0.00
25 T	2-Butanone	0.133	0.124	6.8	107	0.00
26 T	2,2-Dichloropropane	0.695	0.721	-3.7	128	0.00
27 T	cis-1,2-Dichloroethene	0.527	0.542	-2.8	120	0.00
28 T	Bromochloromethane	0.295	0.290	1.7	119	0.00
29 T	Tetrahydrofuran	0.091	0.089	2.2	110	0.00
30 C	Chloroform	0.775	0.794	-2.5#	119	0.00
31 T	Cyclohexane	0.757	0.754	0.4	122	0.00
32 T	1,1,1-Trichloroethane	0.716	0.740	-3.4	122	0.00
33 S	1,2-Dichloroethane-d4	0.391	0.380	2.8	112	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	119	0.00
35 S	Dibromofluoromethane	0.276	0.279	-1.1	118	0.00
36 T	1,1-Dichloropropene	0.423	0.435	-2.8	123	0.00
37 T	Ethyl Acetate	0.208	0.203	2.4	110	0.00
38 T	Carbon Tetrachloride	0.455	0.464	-2.0	120	0.00
39 T	Methylcyclohexane	0.569	0.585	-2.8	122	0.00
40 TM	Benzene	1.235	1.261	-2.1	120	0.00
41 T	Methacrylonitrile	0.125	0.137	-9.6	122	0.00
42 TM	1,2-Dichloroethane	0.326	0.320	1.8	114	0.00
43 T	Isopropyl Acetate	0.387	0.392	-1.3	114	0.00
44 TM	Trichloroethene	0.394	0.402	-2.0	120	0.00
45 C	1,2-Dichloropropane	0.293	0.299	-2.0#	122	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv052020\
 Data File : VY002719.D
 Acq On : 20 May 2020 10:07
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 21 03:18:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y051520S.M
 Quant Title : SW846 8260
 QLast Update : Sat May 16 04:52:36 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.165	0.165	0.0	115	0.00
47 T	Bromodichloromethane	0.402	0.410	-2.0	118	0.00
48 T	Methyl methacrylate	0.178	0.176	1.1	112	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	114	0.00
50 S	Toluene-d8	1.111	1.103	0.7	116	0.00
51 T	4-Methyl-2-Pentanone	0.200	0.198	1.0	112	0.00
52 CM	Toluene	0.799	0.806	-0.9#	119	0.00
53 T	t-1,3-Dichloropropene	0.429	0.436	-1.6	118	0.00
54 T	cis-1,3-Dichloropropene	0.497	0.506	-1.8	118	0.00
55 T	1,1,2-Trichloroethane	0.245	0.245	0.0	117	0.00
56 T	Ethyl methacrylate	0.326	0.330	-1.2	114	0.00
57 T	1,3-Dichloropropane	0.411	0.414	-0.7	116	0.00
58 T	2-Chloroethyl Vinyl ether	0.161	0.159	1.2	112	0.00
59 T	2-Hexanone	0.139	0.135	2.9	108	0.00
60 T	Dibromochloromethane	0.313	0.317	-1.3	115	0.00
61 T	1,2-Dibromoethane	0.242	0.241	0.4	115	0.00
62 S	4-Bromofluorobenzene	0.382	0.367	3.9	115	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	115	0.00
64 T	Tetrachloroethene	0.475	0.493	-3.8	118	0.00
65 PM	Chlorobenzene	0.966	0.991	-2.6	117	0.00
66 T	1,1,1,2-Tetrachloroethane	0.359	0.375	-4.5	118	0.00
67 C	Ethyl Benzene	1.696	1.767	-4.2#	119	0.00
68 T	m/p-Xylenes	0.660	0.687	-4.1	119	0.00
69 T	o-Xylene	0.619	0.644	-4.0	118	0.00
70 T	Styrene	1.064	1.101	-3.5	117	0.00
71 P	Bromoform	0.222	0.229	-3.2	114	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
73 T	Isopropylbenzene	3.267	3.449	-5.6	118	0.00
74 T	N-amyl acetate	0.767	0.793	-3.4	111	0.00
75 P	1,1,2,2-Tetrachloroethane	0.481	0.488	-1.5	112	0.00
76 T	1,2,3-Trichloropropane	0.411	0.411	0.0	108	0.00
77 T	Bromobenzene	0.817	0.850	-4.0	117	0.00
78 T	n-propylbenzene	3.801	4.006	-5.4	118	0.00
79 T	2-Chlorotoluene	2.108	2.196	-4.2	117	0.00
80 T	1,3,5-Trimethylbenzene	2.745	2.851	-3.9	116	0.00
81 T	trans-1,4-Dichloro-2-butene	0.217	0.224	-3.2	114	0.00
82 T	4-Chlorotoluene	2.208	2.266	-2.6	116	0.00
83 T	tert-Butylbenzene	2.430	2.543	-4.7	118	0.00
84 T	1,2,4-Trimethylbenzene	2.744	2.834	-3.3	116	0.00
85 T	sec-Butylbenzene	3.353	3.510	-4.7	118	0.00
86 T	p-Isopropyltoluene	3.138	3.282	-4.6	117	0.00
87 T	1,3-Dichlorobenzene	1.575	1.609	-2.2	116	0.00
88 T	1,4-Dichlorobenzene	1.568	1.582	-0.9	115	0.00
89 T	n-Butylbenzene	2.882	2.995	-3.9	118	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv052020\
Data File : VY002719.D
Acq On : 20 May 2020 10:07
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: May 21 03:18:44 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y051520S.M
Quant Title : SW846 8260
QLast Update : Sat May 16 04:52:36 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.581	0.603	-3.8	116	0.00
91 T	1,2-Dichlorobenzene	1.412	1.415	-0.2	114	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.092	0.091	1.1	109	0.00
93 T	1,2,4-Trichlorobenzene	1.093	1.047	4.2	113	0.00
94 T	Hexachlorobutadiene	0.662	0.645	2.6	116	0.00
95 T	Naphthalene	1.945	1.807	7.1	107	0.00
96 T	1,2,3-Trichlorobenzene	0.947	0.863	8.9	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6