


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv052020\
 Data File : VY002718.D
 Acq On : 20 May 2020 09:31
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y051520S.M
 Title : SW846 8260
 Last Update : Sat May 16 04:52:36 2020



AutoFind: Scans 1761, 1762, 1763; Background Corrected with Scan 1751

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	16756	PASS
75	95	30	60	48.7	51144	PASS
95	95	100	100	100.0	105106	PASS
96	95	5	9	6.9	7282	PASS
173	174	0.00	2	1.2	1145	PASS
174	95	50	100	90.7	95312	PASS
175	174	5	9	7.7	7301	PASS
176	174	95	101	96.7	92144	PASS
177	176	5	9	7.8	7198	PASS