

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052024\
 Data File : VY018323.D
 Acq On : 20 May 2024 10:18
 Operator : SY/MD
 Sample : VY0520SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0520SBL01

Quant Time: May 21 01:20:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	146986	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	288648	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	250611	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	87052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	71940	66.529	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.060%
35) Dibromofluoromethane	7.710	113	85900	56.626	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.260%
50) Toluene-d8	10.179	98	330959	55.364	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.720%
62) 4-Bromofluorobenzene	12.483	95	95352	49.393	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.780%
Target Compounds						
20) Methylene Chloride	4.704	84	14624	9.588	ug/l	Qvalue # 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052024\
Data File : VY018323.D
Acq On : 20 May 2024 10:18
Operator : SY/MD
Sample : VY0520SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0520SBL01

Quant Time: May 21 01:20:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

