

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052024\
 Data File : VY018332.D
 Acq On : 20 May 2024 14:03
 Operator : SY/MD
 Sample : P2518-11
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC-MH-FF

Quant Time: May 21 01:24:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	121453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	244411	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	218891	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	83577	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	70167	78.531	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	157.060%
35) Dibromofluoromethane	7.710	113	72343	56.320	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.640%
50) Toluene-d8	10.179	98	284548	56.216	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.440%
62) 4-Bromofluorobenzene	12.477	95	87246	53.373	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.740%
Target Compounds						
					Qvalue	
16) Acetone	3.948	43	2275	12.759	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052024\
Data File : VY018332.D
Acq On : 20 May 2024 14:03
Operator : SY/MD
Sample : P2518-11
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC-MH-FF

Quant Time: May 21 01:24:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

