

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
 Data File : VY008940.D
 Acq On : 21 May 2022 11:32
 Operator : KP/MD
 Sample : N2919-01
 Misc : 3.72g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P015-SS002-0006-01

Quant Time: May 21 21:28:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	148958	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	263112	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	246479	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	100319	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	89061	58.094	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.180%
35) Dibromofluoromethane	7.716	113	83293	50.908	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.820%
50) Toluene-d8	10.179	98	324698	53.794	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.580%
62) 4-Bromofluorobenzene	12.483	95	94061	38.631	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.260%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.906	85	1426	1.047	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
Data File : VY008940.D
Acq On : 21 May 2022 11:32
Operator : KP/MD
Sample : N2919-01
Misc : 3.72g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P015-SS002-0006-01

Quant Time: May 21 21:28:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

