

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
 Data File : VY008953.D
 Acq On : 21 May 2022 16:37
 Operator : KP/MD
 Sample : N2919-18
 Misc : 6.29g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P015-SS002-0612-01

Quant Time: May 21 21:30:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	120554	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	216552	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	205242	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	90885	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	77400	62.383	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.760%	
35) Dibromofluoromethane	7.716	113	70266	52.180	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.360%	
50) Toluene-d8	10.179	98	272854	54.924	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.840%	
62) 4-Bromofluorobenzene	12.477	95	78241	39.043	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 78.080%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.894	85	1374	1.246	ug/l	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
Data File : VY008953.D
Acq On : 21 May 2022 16:37
Operator : KP/MD
Sample : N2919-18
Misc : 6.29g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P015-SS002-0612-01

Quant Time: May 21 21:30:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

