

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
 Data File : VY008955.D
 Acq On : 21 May 2022 17:23
 Operator : KP/MD
 Sample : N2919-20
 Misc : 5.43g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P015-SS002-1824-01

Quant Time: May 21 21:31:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	120404	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	214952	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	205236	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	88874	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	76509	61.742	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.480%	
35) Dibromofluoromethane	7.716	113	68724	51.415	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.820%	
50) Toluene-d8	10.179	98	269555	54.664	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.320%	
62) 4-Bromofluorobenzene	12.477	95	78924	39.677	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 79.360%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.900	85	1296	1.177	ug/l	81
16) Acetone	3.942	43	4432	17.182	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
Data File : VY008955.D
Acq On : 21 May 2022 17:23
Operator : KP/MD
Sample : N2919-20
Misc : 5.43g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P015-SS002-1824-01

Quant Time: May 21 21:31:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

