

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
 Data File : VY008956.D
 Acq On : 21 May 2022 17:47
 Operator : KP/MD
 Sample : N2919-21
 Misc : 5.50g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P015-SS002-2430-01

Quant Time: May 21 21:31:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	118476	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	215534	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	204551	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	91601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	77197	63.311	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.620%	
35) Dibromofluoromethane	7.716	113	68219	50.899	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.800%	
50) Toluene-d8	10.179	98	266855	53.970	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.940%	
62) 4-Bromofluorobenzene	12.477	95	77896	39.054	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 78.100%	
Target Compounds						
16) Acetone	3.942	43	4602	18.131	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
Data File : VY008956.D
Acq On : 21 May 2022 17:47
Operator : KP/MD
Sample : N2919-21
Misc : 5.50g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P015-SS002-2430-01

Quant Time: May 21 21:31:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

