

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
 Data File : VY008961.D
 Acq On : 21 May 2022 19:44
 Operator : KP/MD
 Sample : N2888-09
 Misc : 5.57g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P042-SS001-2430-01

Quant Time: May 21 21:32:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	112875	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	206192	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	195749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	87270	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	77915	67.070	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 134.140%	
35) Dibromofluoromethane	7.710	113	66475	51.845	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.680%	
50) Toluene-d8	10.179	98	258591	54.668	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.340%	
62) 4-Bromofluorobenzene	12.477	95	75985	39.822	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 79.640%	
Target Compounds						
70) Styrene	12.038	104	7137	1.360	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
Data File : VY008961.D
Acq On : 21 May 2022 19:44
Operator : KP/MD
Sample : N2888-09
Misc : 5.57g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P042-SS001-2430-01

Quant Time: May 21 21:32:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

