

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
 Data File : VY008973.D
 Acq On : 22 May 2022 01:54
 Operator : KP/MD
 Sample : N2887-03
 Misc : 5.53g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P030-SS003-0006-01

Quant Time: May 22 02:42:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	118590	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	215187	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	206903	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	88198	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	78554	64.362	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.720%
35) Dibromofluoromethane	7.716	113	69316	51.801	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.600%
50) Toluene-d8	10.179	98	268391	54.368	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.740%
62) 4-Bromofluorobenzene	12.477	95	76998	38.666	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
Data File : VY008973.D
Acq On : 22 May 2022 01:54
Operator : KP/MD
Sample : N2887-03
Misc : 5.53g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P030-SS003-0006-01

Quant Time: May 22 02:42:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

