

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
 Data File : VY008975.D
 Acq On : 22 May 2022 02:41
 Operator : KP/MD
 Sample : N2887-08
 Misc : 5.95g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P038-SS001-0612-01

Quant Time: May 22 04:48:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	116101	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	209558	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	199950	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	85187	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	78365	65.583	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.160%
35) Dibromofluoromethane	7.716	113	68464	52.539	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.080%
50) Toluene-d8	10.179	98	265201	55.165	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.340%
62) 4-Bromofluorobenzene	12.477	95	75556	38.961	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
Data File : VY008975.D
Acq On : 22 May 2022 02:41
Operator : KP/MD
Sample : N2887-08
Misc : 5.95g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P038-SS001-0612-01

Quant Time: May 22 04:48:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

