

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
 Data File : VY008987.D
 Acq On : 22 May 2022 07:22
 Operator : KP/MD
 Sample : N2887-20
 Misc : 5.95g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P040-SS001-1824-01

Quant Time: May 23 02:27:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	113069	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	204274	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	199008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	85699	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	74903	64.367	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.740%	
35) Dibromofluoromethane	7.716	113	66443	52.307	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.620%	
50) Toluene-d8	10.179	98	261470	55.796	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 111.600%	
62) 4-Bromofluorobenzene	12.477	95	77562	41.030	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 82.060%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
Data File : VY008987.D
Acq On : 22 May 2022 07:22
Operator : KP/MD
Sample : N2887-20
Misc : 5.95g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P040-SS001-1824-01

Quant Time: May 23 02:27:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

