

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
 Data File : VY008988.D
 Acq On : 22 May 2022 07:45
 Operator : KP/MD
 Sample : N2887-21
 Misc : 4.83g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P040-SS001-2430-01

Quant Time: May 23 02:28:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	117803	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	216048	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	207265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	90916	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	79808	65.826	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.660%
35) Dibromofluoromethane	7.716	113	69320	51.597	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.200%
50) Toluene-d8	10.179	98	275830	55.653	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.300%
62) 4-Bromofluorobenzene	12.483	95	78920	39.474	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
Data File : VY008988.D
Acq On : 22 May 2022 07:45
Operator : KP/MD
Sample : N2887-21
Misc : 4.83g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P040-SS001-2430-01

Quant Time: May 23 02:28:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

