

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
 Data File : VY008990.D
 Acq On : 22 May 2022 08:32
 Operator : KP/MD
 Sample : N2919-05
 Misc : 4.96g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P028-SS001-0612-01

Quant Time: May 23 02:28:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	112129	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	203872	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	199007	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	87800	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	74727	64.754	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.500%
35) Dibromofluoromethane	7.716	113	65995	52.056	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.120%
50) Toluene-d8	10.179	98	263575	56.356	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.720%
62) 4-Bromofluorobenzene	12.477	95	79748	42.270	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
Data File : VY008990.D
Acq On : 22 May 2022 08:32
Operator : KP/MD
Sample : N2919-05
Misc : 4.96g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P028-SS001-0612-01

Quant Time: May 23 02:28:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

