

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
 Data File : VY008993.D
 Acq On : 22 May 2022 09:40
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 23 02:29:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	75	0.00
2 T	Dichlorodifluoromethane	0.457	0.490	-7.2	76	0.00
3 P	Chloromethane	0.655	0.839	-28.1#	91	0.00
4 C	Vinyl Chloride	0.971	1.161	-19.6#	86	0.00
5 T	Bromomethane	0.951	0.990	-4.1	90	0.00
6 T	Chloroethane	0.688	0.836	-21.5	92	0.00
7 T	Trichlorofluoromethane	0.824	0.865	-5.0	77	0.00
8 T	Diethyl Ether	0.258	0.298	-15.5	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.507	0.515	-1.6	74	0.00
10 T	Methyl Iodide	0.523	0.596	-14.0	73	0.00
11 T	Tert butyl alcohol	0.052	0.054	-3.8	80	0.00
12 CM	1,1-Dichloroethene	0.470	0.505	-7.4#	79	0.00
13 T	Acrolein	0.017	0.009	47.1#	73	0.00
14 T	Allyl chloride	0.709	0.777	-9.6	79	0.00
15 T	Acrylonitrile	0.144	0.169	-17.4	88	0.00
16 T	Acetone	0.107	0.124	-15.9	89	0.00
17 T	Carbon Disulfide	1.536	1.685	-9.7	79	0.00
18 T	Methyl Acetate	0.334	0.509	-52.4#	113	0.00
19 T	Methyl tert-butyl Ether	1.412	1.631	-15.5	85	0.00
20 T	Methylene Chloride	0.788	0.643	18.4	78	0.00
21 T	trans-1,2-Dichloroethene	0.604	0.655	-8.4	82	0.00
22 T	Diisopropyl ether	1.703	2.003	-17.6	86	0.00
23 T	Vinyl Acetate	1.126	1.272	-13.0	80	0.00
24 P	1,1-Dichloroethane	1.018	1.156	-13.6	84	0.00
25 T	2-Butanone	0.191	0.230	-20.4	87	0.00
26 T	2,2-Dichloropropane	0.938	0.736	21.5	58	0.00
27 T	cis-1,2-Dichloroethene	0.680	0.753	-10.7	80	0.00
28 T	Bromochloromethane	0.350	0.409	-16.9	90	0.00
29 T	Tetrahydrofuran	0.130	0.158	-21.5	87	0.00
30 C	Chloroform	1.099	1.235	-12.4#	83	0.00
31 T	Cyclohexane	1.007	1.001	0.6	75	0.00
32 T	1,1,1-Trichloroethane	0.997	1.089	-9.2	79	0.00
33 S	1,2-Dichloroethane-d4	0.515	0.470	8.7	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.311	0.279	10.3	86	0.00
36 T	1,1-Dichloropropene	0.529	0.527	0.4	78	0.00
37 T	Ethyl Acetate	0.274	0.288	-5.1	83	0.00
38 T	Carbon Tetrachloride	0.554	0.537	3.1	76	0.00
39 T	Methylcyclohexane	0.683	0.643	5.9	71	0.00
40 TM	Benzene	1.526	1.617	-6.0	82	0.00
41 T	Methacrylonitrile	0.142	0.165	-16.2	88	0.00
42 TM	1,2-Dichloroethane	0.448	0.483	-7.8	84	0.00
43 T	Isopropyl Acetate	0.510	0.571	-12.0	86	0.00
44 TM	Trichloroethene	0.404	0.408	-1.0	79	0.00
45 C	1,2-Dichloropropane	0.366	0.395	-7.9#	83	0.00
46 T	Dibromomethane	0.233	0.250	-7.3	83	0.00
47 T	Bromodichloromethane	0.532	0.564	-6.0	83	0.00
48 T	Methyl methacrylate	0.228	0.262	-14.9	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
 Data File : VY008993.D
 Acq On : 22 May 2022 09:40
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 23 02:29:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	80	0.00
50 S	Toluene-d8	1.147	0.934	18.6	85	0.00
51 T	4-Methyl-2-Pentanone	0.282	0.330	-17.0	87	0.00
52 CM	Toluene	1.066	1.155	-8.3#	81	0.00
53 T	t-1,3-Dichloropropene	0.565	0.578	-2.3	76	0.00
54 T	cis-1,3-Dichloropropene	0.608	0.598	1.6	76	0.00
55 T	1,1,2-Trichloroethane	0.326	0.355	-8.9	83	0.00
56 T	Ethyl methacrylate	0.417	0.483	-15.8	84	0.00
57 T	1,3-Dichloropropane	0.532	0.591	-11.1	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.180	0.192	-6.7	80	0.00
59 T	2-Hexanone	0.193	0.229	-18.7	85	0.00
60 T	Dibromochloromethane	0.395	0.425	-7.6	81	0.00
61 T	1,2-Dibromoethane	0.323	0.358	-10.8	83	0.00
62 S	4-Bromofluorobenzene	0.463	0.483	-4.3	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.364	0.356	2.2	77	0.00
65 PM	Chlorobenzene	1.148	1.178	-2.6	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.420	0.444	-5.7	81	0.00
67 C	Ethyl Benzene	2.068	2.172	-5.0#	78	0.00
68 T	m/p-Xylenes	0.840	0.915	-8.9	78	0.00
69 T	o-Xylene	0.780	0.864	-10.8	80	0.00
70 T	Styrene	1.341	1.542	-15.0	79	0.00
71 P	Bromoform	0.279	0.309	-10.8	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
73 T	Isopropylbenzene	3.739	3.523	5.8	77	0.00
74 T	N-amyl acetate	0.863	0.872	-1.0	81	0.00
75 P	1,1,2,2-Tetrachloroethane	0.770	0.740	3.9	81	0.00
76 T	1,2,3-Trichloropropane	0.537	0.522	2.8	77	0.00
77 T	Bromobenzene	0.867	0.819	5.5	79	0.00
78 T	n-propylbenzene	4.320	4.397	-1.8	122	0.00
79 T	2-Chlorotoluene	2.492	2.483	0.4	79	0.00
80 T	1,3,5-Trimethylbenzene	3.134	3.128	0.2	77	0.00
81 T	trans-1,4-Dichloro-2-butene	0.262	0.210	19.8	66	0.00
82 T	4-Chlorotoluene	2.665	2.711	-1.7	78	0.00
83 T	tert-Butylbenzene	2.780	2.728	1.9	77	0.00
84 T	1,2,4-Trimethylbenzene	3.111	3.227	-3.7	79	0.00
85 T	sec-Butylbenzene	4.254	4.261	-0.2	76	0.00
86 T	p-Isopropyltoluene	3.547	3.611	-1.8	77	0.00
87 T	1,3-Dichlorobenzene	1.849	1.863	-0.8	79	0.00
88 T	1,4-Dichlorobenzene	1.857	1.853	0.2	78	0.00
89 T	n-Butylbenzene	3.245	3.154	2.8	75	0.00
90 T	Hexachloroethane	0.665	0.639	3.9	77	0.00
91 T	1,2-Dichlorobenzene	1.632	1.668	-2.2	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.112	6.7	77	0.00
93 T	1,2,4-Trichlorobenzene	0.896	0.800	10.7	75	0.00
94 T	Hexachlorobutadiene	0.488	0.420	13.9	73	0.00
95 T	Naphthalene	1.799	1.774	1.4	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052122\
Data File : VY008993.D
Acq On : 22 May 2022 09:40
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 60 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: May 23 02:29:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.778	0.725	6.8	79	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6