


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY052220\
 Data File : VY002746.D
 Acq On : 22 May 2020 09:21
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y051520S.M
 Title : SW846 8260
 Last Update : Sat May 16 04:52:36 2020



AutoFind: Scans 1761, 1762, 1763; Background Corrected with Scan 1753

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.2	13617	PASS
75	95	30	60	48.0	40320	PASS
95	95	100	100	100.0	84029	PASS
96	95	5	9	6.8	5677	PASS
173	174	0.00	2	1.0	787	PASS
174	95	50	100	90.4	75984	PASS
175	174	5	9	7.9	6017	PASS
176	174	95	101	97.5	74082	PASS
177	176	5	9	7.7	5708	PASS